

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120519\
 Data File : VX013781.D
 Acq On : 05 Dec 2019 17:37
 Operator : JC/SP
 Sample : K6178-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 391203051.2

Quant Time: Dec 06 03:21:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 12 01:16:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	103446	30.00	ug/l	0.01
28) 1,4-Difluorobenzene	6.85	114	565011	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	492953	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.05	65	214175	28.04	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	93.47%
60) 4-Bromofluorobenzene	11.14	95	224361	28.32	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	94.40%
63) Toluene-d8	8.71	98	666615	29.90	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	99.67%

Target Compounds

					Ovalue
25) Chloroform	5.20	83	102579	8.500 ug/l	92
41) Bromodichloromethane	7.89	83	31692	3.604 ug/l	95
53) Dibromochloromethane	9.57	129	11722	1.659 ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120519\
Data File : VX013781.D
Acq On : 05 Dec 2019 17:37
Operator : JC/SP
Sample : K6178-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
391203051.2

Quant Time: Dec 06 03:21:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 12 01:16:13 2019
Response via : Initial Calibration

