

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120522\
 Data File : VX033230.D
 Acq On : 05 Dec 2022 17:00
 Operator : JC/MD
 Sample : N5852-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB03-12012022

Quant Time: Dec 06 06:00:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	101955	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	166360	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	150965	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68205	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	52922	53.330	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	106.660%	
35) Dibromofluoromethane	5.385	113	44486	47.433	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	94.860%	
50) Toluene-d8	8.647	98	141318	50.083	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	100.160%	
62) 4-Bromofluorobenzene	11.079	95	52323	42.470	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	84.940%	
Target Compounds						
						Qvalue
16) Acetone	2.398	43	1128	2.433	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120522\
Data File : VX033230.D
Acq On : 05 Dec 2022 17:00
Operator : JC/MD
Sample : N5852-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB03-12012022

Quant Time: Dec 06 06:00:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

