

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120522\
 Data File : VX033234.D
 Acq On : 05 Dec 2022 18:32
 Operator : JC/MD
 Sample : PB149407ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149407ZHE#09

Quant Time: Dec 06 06:00:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	105405	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	176350	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	168145	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74621	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56985	55.683	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.360%
35) Dibromofluoromethane	5.385	113	47497	47.774	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.540%
50) Toluene-d8	8.647	98	156051	52.265	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.540%
62) 4-Bromofluorobenzene	11.079	95	60878	46.615	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.220%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	10605	22.122	ug/l	96
20) Methylene Chloride	2.794	84	3319	2.084	ug/l	96
37) Ethyl Acetate	4.715	43	15414	11.114	ug/l	99
43) Isopropyl Acetate	6.336	43	73439	32.018	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120522\
Data File : VX033234.D
Acq On : 05 Dec 2022 18:32
Operator : JC/MD
Sample : PB149407ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149407ZHE#09

Quant Time: Dec 06 06:00:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

