

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120522\
 Data File : VX033249.D
 Acq On : 06 Dec 2022 00:17
 Operator : JC/MD
 Sample : PB149407ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149407ZHE#07

Quant Time: Dec 06 06:03:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126895	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	207983	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	191505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85929	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	35102	26.859	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	53.720%#
35) Dibromofluoromethane	5.391	113	38409	32.757	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	65.520%#
50) Toluene-d8	8.647	98	81456	21.875	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	43.760%#
62) 4-Bromofluorobenzene	11.079	95	62750	40.740	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	81.480%#
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	11498	19.923	ug/l	99
18) Methyl Acetate	2.703	43	1652	1.034	ug/l #	89
20) Methylene Chloride	2.794	84	4006	2.093	ug/l	93
37) Ethyl Acetate	4.721	43	18239	11.151	ug/l	99
43) Isopropyl Acetate	6.336	43	83934	31.028	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120522\
Data File : VX033249.D
Acq On : 06 Dec 2022 00:17
Operator : JC/MD
Sample : PB149407ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149407ZHE#07

Quant Time: Dec 06 06:03:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

