

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120522\
 Data File : VX033251.D
 Acq On : 06 Dec 2022 01:03
 Operator : JC/MD
 Sample : PB149407TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149407TB

Quant Time: Dec 06 06:04:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	124911	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	203396	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185864	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85700	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64444	52.985	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.980%			
35) Dibromofluoromethane	5.391	113	53454	46.617	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.240%			
50) Toluene-d8	8.647	98	176460	51.198	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 102.400%			
62) 4-Bromofluorobenzene	11.079	95	68215	45.287	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 90.580%			
Target Compounds						
16) Acetone	2.386	43	11677	20.554	ug/l	95
20) Methylene Chloride	2.794	84	4060	2.194	ug/l #	88
37) Ethyl Acetate	4.714	43	17578	10.989	ug/l	99
43) Isopropyl Acetate	6.342	43	80483	30.423	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120522\
Data File : VX033251.D
Acq On : 06 Dec 2022 01:03
Operator : JC/MD
Sample : PB149407TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149407TB

Quant Time: Dec 06 06:04:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

