

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120522\
 Data File : VX033225.D
 Acq On : 05 Dec 2022 14:50
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 06 05:59:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.382	0.306	19.9	78	0.00
3 P	Chloromethane	0.412	0.337	18.2	81	0.00
4 C	Vinyl Chloride	0.459	0.396	13.7#	83	0.00
5 T	Bromomethane	0.340	0.375	-10.3	110	0.01
6 T	Chloroethane	0.374	0.369	1.3	91	0.00
7 T	Trichlorofluoromethane	0.891	0.881	1.1	95	0.00
8 T	Diethyl Ether	0.281	0.297	-5.7	101	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.468	0.467	0.2	93	0.00
10 T	Methyl Iodide	0.536	0.458	14.6	74	0.00
11 T	Tert butyl alcohol	0.077	0.068	11.7	82	0.02
12 CM	1,1-Dichloroethene	0.436	0.421	3.4#	90	0.00
13 T	Acrolein	0.089	0.111	-24.7	122	0.00
14 T	Allyl chloride	0.704	0.691	1.8	91	0.00
15 T	Acrylonitrile	0.218	0.230	-5.5	100	0.00
16 T	Acetone	0.227	0.215	5.3	88	0.00
17 T	Carbon Disulfide	1.205	1.019	15.4	83	0.00
18 T	Methyl Acetate	0.629	0.656	-4.3	100	0.00
19 T	Methyl tert-butyl Ether	1.513	1.536	-1.5	94	0.00
20 T	Methylene Chloride	0.548	0.476	13.1	91	0.00
21 T	trans-1,2-Dichloroethene	0.499	0.466	6.6	89	0.00
22 T	Diisopropyl ether	1.534	1.639	-6.8	95	0.00
23 T	Vinyl Acetate	1.201	1.219	-1.5	90	0.00
24 P	1,1-Dichloroethane	0.878	0.865	1.5	93	0.00
25 T	2-Butanone	0.314	0.333	-6.1	99	0.00
26 T	2,2-Dichloropropane	0.706	0.712	-0.8	91	0.00
27 T	cis-1,2-Dichloroethene	0.567	0.579	-2.1	93	0.00
28 T	Bromochloromethane	0.386	0.371	3.9	88	0.00
29 T	Tetrahydrofuran	0.199	0.216	-8.5	103	0.00
30 C	Chloroform	0.983	1.035	-5.3#	96	0.00
31 T	Cyclohexane	0.789	0.767	2.8	89	0.00
32 T	1,1,1-Trichloroethane	0.880	0.929	-5.6	95	0.00
33 S	1,2-Dichloroethane-d4	0.539	0.423	21.5	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.282	0.244	13.5	85	0.00
36 T	1,1-Dichloropropene	0.451	0.453	-0.4	92	0.00
37 T	Ethyl Acetate	0.393	0.416	-5.9	97	0.00
38 T	Carbon Tetrachloride	0.500	0.537	-7.4	98	0.00
39 T	Methylcyclohexane	0.514	0.524	-1.9	93	0.00
40 TM	Benzene	1.264	1.310	-3.6	95	0.00
41 T	Methacrylonitrile	0.221	0.242	-9.5	100	0.00
42 TM	1,2-Dichloroethane	0.509	0.523	-2.8	93	0.00
43 T	Isopropyl Acetate	0.650	0.701	-7.8	98	0.00
44 TM	Trichloroethene	0.355	0.360	-1.4	92	0.00
45 C	1,2-Dichloropropane	0.312	0.335	-7.4#	96	0.00
46 T	Dibromomethane	0.244	0.253	-3.7	96	0.00
47 T	Bromodichloromethane	0.471	0.531	-12.7	100	0.00
48 T	Methyl methacrylate	0.335	0.353	-5.4	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120522\
 Data File : VX033225.D
 Acq On : 05 Dec 2022 14:50
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 06 05:59:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.007	-16.7	104	0.02
50 S	Toluene-d8	0.915	0.716	21.7	80	0.00
51 T	4-Methyl-2-Pentanone	0.401	0.458	-14.2	103	0.00
52 CM	Toluene	0.831	0.878	-5.7#	95	0.00
53 T	t-1,3-Dichloropropene	0.462	0.540	-16.9	98	0.00
54 T	cis-1,3-Dichloropropene	0.508	0.575	-13.2	99	0.00
55 T	1,1,2-Trichloroethane	0.338	0.371	-9.8	101	0.00
56 T	Ethyl methacrylate	0.466	0.517	-10.9	95	0.00
57 T	1,3-Dichloropropane	0.546	0.592	-8.4	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.221	0.243	-10.0	99	0.00
59 T	2-Hexanone	0.296	0.337	-13.9	100	0.00
60 T	Dibromochloromethane	0.364	0.442	-21.4	104	0.00
61 T	1,2-Dibromoethane	0.357	0.398	-11.5	99	0.00
62 S	4-Bromofluorobenzene	0.370	0.326	11.9	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.354	0.338	4.5	92	0.00
65 PM	Chlorobenzene	0.997	1.030	-3.3	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.365	0.414	-13.4	103	0.00
67 C	Ethyl Benzene	1.768	1.871	-5.8#	95	0.00
68 T	m/p-Xylenes	0.692	0.738	-6.6	97	0.00
69 T	o-Xylene	0.670	0.725	-8.2	97	0.00
70 T	Styrene	1.096	1.231	-12.3	98	0.00
71 P	Bromoform	0.280	0.350	-25.0	110	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.573	3.748	-4.9	97	0.00
74 T	N-amyl acetate	1.215	1.277	-5.1	92	0.00
75 P	1,1,2,2-Tetrachloroethane	1.078	1.160	-7.6	102	0.00
76 T	1,2,3-Trichloropropane	0.947	1.046	-10.5	103	0.00
77 T	Bromobenzene	0.879	0.907	-3.2	97	0.00
78 T	n-propylbenzene	4.085	4.363	-6.8	97	0.00
79 T	2-Chlorotoluene	2.481	2.574	-3.7	97	0.00
80 T	1,3,5-Trimethylbenzene	3.002	3.192	-6.3	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.291	0.350	-20.3	108	0.00
82 T	4-Chlorotoluene	2.908	3.002	-3.2	95	0.00
83 T	tert-Butylbenzene	2.988	3.215	-7.6	98	0.00
84 T	1,2,4-Trimethylbenzene	2.993	3.155	-5.4	95	0.00
85 T	sec-Butylbenzene	3.574	3.884	-8.7	98	0.00
86 T	p-Isopropyltoluene	3.028	3.316	-9.5	97	0.00
87 T	1,3-Dichlorobenzene	1.648	1.724	-4.6	98	0.00
88 T	1,4-Dichlorobenzene	1.688	1.750	-3.7	97	0.00
89 T	n-Butylbenzene	2.609	2.846	-9.1	97	0.00
90 T	Hexachloroethane	0.498	0.609	-22.3	112	0.00
91 T	1,2-Dichlorobenzene	1.603	1.700	-6.1	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.215	0.249	-15.8	103	0.00
93 T	1,2,4-Trichlorobenzene	0.988	1.059	-7.2	99	0.00
94 T	Hexachlorobutadiene	0.436	0.430	1.4	96	0.00
95 T	Naphthalene	3.197	3.503	-9.6	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120522\
Data File : VX033225.D
Acq On : 05 Dec 2022 14:50
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 06 05:59:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.991	1.064	-7.4	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6