

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120524\
 Data File : VX044154.D
 Acq On : 05 Dec 2024 17:49
 Operator : JC/MD
 Sample : P5066-11ME
 Misc : 4.48g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 E28M9ME

Quant Time: Dec 06 04:10:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120524WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 06 03:51:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	239877	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	211576	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	89325	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	88658	40.388	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.780%
7) Chloroethane-d5	1.642	69	66717	39.235	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.460%
11) 1,1-Dichloroethene-d2	2.294	65	36448	38.824	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.640%
21) 2-Butanone-d5	4.465	46	121977	90.758	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.760%
24) Chloroform-d	5.050	84	170987	41.025	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.040%
26) 1,2-Dichloroethane-d4	5.952	65	127556	43.276	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.560%
32) Benzene-d6	5.964	84	320169	43.795	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.600%
36) 1,2-Dichloropropane-d6	7.306	67	101869	44.876	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.760%
41) Toluene-d8	8.647	98	287260	42.497	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.000%
43) trans-1,3-Dichloroprop...	8.952	79	43869	40.215	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.420%
47) 2-Hexanone-d5	9.384	63	83119	93.910	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.910%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	136646	44.893	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.780%
66) 1,2-Dichlorobenzene-d4	12.317	152	95386	45.831	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.660%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120524\
Data File : VX044154.D
Acq On : 05 Dec 2024 17:49
Operator : JC/MD
Sample : P5066-11ME
Misc : 4.48g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
E28M9ME

Quant Time: Dec 06 04:10:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120524WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 06 03:51:43 2024
Response via : Initial Calibration

