

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120525\
 Data File : VX048731.D
 Acq On : 05 Dec 2025 13:32
 Operator : JC/MD
 Sample : Q3757-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14A-17.5-120325

Quant Time: Dec 05 22:03:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	176619	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.744	114	302604	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	274042	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	141373	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	124177	52.383	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.760%
35) Dibromofluoromethane	5.367	113	103142	49.503	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.000%
50) Toluene-d8	8.628	98	355343	48.654	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.300%
62) 4-Bromofluorobenzene	11.061	95	125871	49.979	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.960%
Target Compounds						Qvalue
16) Acetone	2.373	43	3056	3.368	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120525\
Data File : VX048731.D
Acq On : 05 Dec 2025 13:32
Operator : JC/MD
Sample : Q3757-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14A-17.5-120325

Quant Time: Dec 05 22:03:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

