

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120525\
 Data File : VX048738.D
 Acq On : 05 Dec 2025 15:55
 Operator : JC/MD
 Sample : Q3789-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 05 22:06:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	140193	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	290551	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	224650	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	129713	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	128233	68.149	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	136.300%#
35) Dibromofluoromethane	5.373	113	87231	43.603	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.200%
50) Toluene-d8	8.635	98	293644	41.874	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	83.740%#
62) 4-Bromofluorobenzene	11.061	95	104604	43.258	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120525\
Data File : VX048738.D
Acq On : 05 Dec 2025 15:55
Operator : JC/MD
Sample : Q3789-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 05 22:06:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

