



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120525\
Data File : VX048722.D
Acq On : 05 Dec 2025 08:14
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Title : SW846 8260
Last Update : Fri Dec 05 03:27:34 2025



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1629

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.3	5123	PASS
75	95	30	60	48.5	15274	PASS
95	95	100	100	100.0	31461	PASS
96	95	5	9	6.6	2062	PASS
173	174	0.00	2	1.2	294	PASS
174	95	50	100	80.6	25363	PASS
175	174	5	9	7.3	1854	PASS
176	174	95	101	96.4	24445	PASS
177	176	5	9	7.0	1707	PASS