

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120618\
 Data File : VX006309.D
 Acq On : 06 Dec 2018 11:12
 Operator : JC/MD
 Sample : VX1206WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1206WBL01

Quant Time: Dec 07 04:40:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	751674	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1124211	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1007609	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	497326	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	364144	47.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.92%	
35) Dibromofluoromethane	5.50	113	351334	48.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.54%	
50) Toluene-d8	8.71	98	1291018	48.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.52%	
62) 4-Bromofluorobenzene	11.13	95	482102	48.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.64%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120618\
Data File : VX006309.D
Acq On : 06 Dec 2018 11:12
Operator : JC/MD
Sample : VX1206WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1206WBL01

Quant Time: Dec 07 04:40:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

