

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120618\
 Data File : VX006310.D
 Acq On : 06 Dec 2018 11:38
 Operator : JC/MD
 Sample : VX1206MBL01
 Misc : 5.0g/10mL/100uL/5.0ml/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1206MBL01

Quant Time: Dec 07 04:42:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	868108	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	1317666	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1197378	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	582268	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	430481	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
35) Dibromofluoromethane	5.51	113	415188	48.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.34%	
50) Toluene-d8	8.71	98	1536928	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
62) 4-Bromofluorobenzene	11.13	95	570817	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120618\
Data File : VX006310.D
Acq On : 06 Dec 2018 11:38
Operator : JC/MD
Sample : VX1206MBL01
Misc : 5.0g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1206MBL01

Quant Time: Dec 07 04:42:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

