

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120618\
 Data File : VX006316.D
 Acq On : 06 Dec 2018 14:18
 Operator : JC/MD
 Sample : J6250-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW30411-20181204

Quant Time: Dec 07 05:06:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	803720	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1217220	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1110531	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	544880	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	402645	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
35) Dibromofluoromethane	5.50	113	382507	48.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.08%	
50) Toluene-d8	8.71	98	1422128	49.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.22%	
62) 4-Bromofluorobenzene	11.13	95	530923	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
Target Compounds						
64) Tetrachloroethene	9.34	164	4048	0.484	ug/l #	Qvalue 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120618\
Data File : VX006316.D
Acq On : 06 Dec 2018 14:18
Operator : JC/MD
Sample : J6250-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW30411-20181204

Quant Time: Dec 07 05:06:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

