

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120618\
 Data File : VX006322.D
 Acq On : 06 Dec 2018 16:58
 Operator : JC/MD
 Sample : J6250-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW30412-20181204

Quant Time: Dec 07 05:46:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	794037	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1210825	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1107875	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	526525	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	401581	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
35) Dibromofluoromethane	5.51	113	376585	48.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.08%	
50) Toluene-d8	8.71	98	1405502	49.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.58%	
62) 4-Bromofluorobenzene	11.13	95	522996	48.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.34%	
Target Compounds						
64) Tetrachloroethene	9.34	164	2867	0.344	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120618\
Data File : VX006322.D
Acq On : 06 Dec 2018 16:58
Operator : JC/MD
Sample : J6250-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW304I2-20181204

Quant Time: Dec 07 05:46:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

