

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120618\  
 Data File : VX006323.D  
 Acq On : 06 Dec 2018 17:24  
 Operator : JC/MD  
 Sample : J6259-18  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BPS1-TT-MW302S-20181205

Quant Time: Dec 07 05:48:50 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X112918W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Nov 30 03:55:33 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units  | Dev(Min) |
|-----------------------------|-------|------|----------|-------|--------|----------|
| 1) Pentafluorobenzene       | 5.67  | 168  | 757559   | 50.00 | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.86  | 114  | 1173341  | 50.00 | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.12 | 117  | 1062221  | 50.00 | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.08 | 152  | 508701   | 50.00 | ug/l   | 0.00     |
| System Monitoring Compounds |       |      |          |       |        |          |
| 33) 1,2-Dichloroethane-d4   | 6.07  | 65   | 382647   | 49.48 | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 98.96% |          |
| 35) Dibromofluoromethane    | 5.50  | 113  | 363377   | 47.83 | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 95.66% |          |
| 50) Toluene-d8              | 8.71  | 98   | 1361882  | 49.28 | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 98.56% |          |
| 62) 4-Bromofluorobenzene    | 11.13 | 95   | 503776   | 48.38 | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 96.76% |          |
| Target Compounds            |       |      |          |       |        |          |
| 44) Trichloroethene         | 7.21  | 130  | 4070     | 0.454 | ug/l   | 98       |
| 64) Tetrachloroethene       | 9.34  | 164  | 5832     | 0.729 | ug/l   | 88       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120618\  
Data File : VX006323.D  
Acq On : 06 Dec 2018 17:24  
Operator : JC/MD  
Sample : J6259-18  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BPS1-TT-MW302S-20181205

Quant Time: Dec 07 05:48:50 2018  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X112918W.M  
Quant Title : SW846 8260  
QLast Update : Fri Nov 30 03:55:33 2018  
Response via : Initial Calibration

