

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120618\
 Data File : VX006324.D
 Acq On : 06 Dec 2018 17:51
 Operator : JC/MD
 Sample : J6259-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW306S-20181204

Quant Time: Dec 07 05:50:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	764961	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1181477	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1079100	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	518033	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	390350	49.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.98%	
35) Dibromofluoromethane	5.51	113	371273	48.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.06%	
50) Toluene-d8	8.71	98	1382801	49.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.40%	
62) 4-Bromofluorobenzene	11.13	95	511198	48.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.50%	
Target Compounds						
44) Trichloroethene	7.22	130	5059	0.560	ug/l	93
64) Tetrachloroethene	9.34	164	3547	0.436	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120618\
Data File : VX006324.D
Acq On : 06 Dec 2018 17:51
Operator : JC/MD
Sample : J6259-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW306S-20181204

Quant Time: Dec 07 05:50:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

