

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120618\
 Data File : VX006319.D
 Acq On : 06 Dec 2018 15:38
 Operator : JC/MD
 Sample : J6250-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW307S-20181204

Quant Time: Dec 07 05:42:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	785666	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1209440	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1094333	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	532070	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	393944	49.12	ug/l	0.00
Spiked Amount			Recovery	=	98.24%	
35) Dibromofluoromethane	5.51	113	373848	47.74	ug/l	0.00
Spiked Amount			Recovery	=	95.48%	
50) Toluene-d8	8.71	98	1397950	49.08	ug/l	0.00
Spiked Amount			Recovery	=	98.16%	
62) 4-Bromofluorobenzene	11.13	95	519471	48.40	ug/l	0.00
Spiked Amount			Recovery	=	96.80%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120618\
Data File : VX006319.D
Acq On : 06 Dec 2018 15:38
Operator : JC/MD
Sample : J6250-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW307S-20181204

Quant Time: Dec 07 05:42:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.00 min

Lab File: VX006319.D

Acq: 06 Dec 2018 15:38

Instrument :

MSVOA_X

ClientSampleId :


BPS1-TT-MW307S-20181204

Tot Ion:168 Resp: 785666

Ion Ratio Lower Upper

168 100

99 49.4 37.8 56.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

300000

250000

200000



150000

100000

50000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 49.118 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.00 min

Lab File: VX006319.D


Acq: 06 Dec 2018 15:38

Tgt Ion: 65 Resp: 393944

Ion Ratio Lower Upper

65 100

67 53.0 0.0 107.2



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07


150000


100000

50000

0

Time--> 5.90 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. 0.00 min

Lab File: VX006319.D

Acq: 06 Dec 2018 15:38

Instrument :

MSVOA_X

ClientSampleId :

BPS1-TT-MW307S-20181204


Tot Ion:114 Resp: 1209440

Ion Ratio Lower Upper

114 100

63 17.2 0.0 36.6




88 14.8 0.0 29.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 47.742 ug/l

RT: 5.51 min Scan# 714

Delta R.T. 0.01 min

Lab File: VX006319.D

Acq: 06 Dec 2018 15:38


Tgt Ion:113 Resp: 373848

Ion Ratio Lower Upper

113 100

111 102.6 81.1 121.7



192 22.1 18.2 27.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 49.080 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. 0.00 min
Lab File: VX006319.D
Acq: 06 Dec 2018 15:38




Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW307S-20181204


Tot Ion: 98 Resp: 1397950
Ion Ratio Lower Upper
98 100
100 64.7 52.2 78.2



#62
4-Bromofluorobenzene
Concen: 48.398 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. -0.01 min
Lab File: VX006319.D
Acq: 06 Dec 2018 15:38

Tgt Ion: 95 Resp: 519471
Ion Ratio Lower Upper
95 100
174 90.1 0.0 187.0
176 86.3 0.0 181.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. 0.00 min
Lab File: VX006319.D
Acq: 06 Dec 2018 15:38

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW307S-20181204

Tot Ion:117 Resp: 1094333
Ion Ratio Lower Upper
117 100
82 50.0 39.6 59.4
119 32.6 26.3 39.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX006319.D
Acq: 06 Dec 2018 15:38

Tgt Ion:152 Resp: 532070
Ion Ratio Lower Upper
152 100
115 56.4 39.0 116.9
150 155.5 0.0 345.8

