

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120619\
 Data File : VX013789.D
 Acq On : 06 Dec 2019 11:11
 Operator : JC/SP
 Sample : K6099-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OF-3

Quant Time: Dec 07 03:28:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	668923	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1012456	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	910450	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	420627	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	366773	47.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.48%	
35) Dibromofluoromethane	5.48	113	303913	48.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.12%	
50) Toluene-d8	8.71	98	1191543	48.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.72%	
62) 4-Bromofluorobenzene	11.13	95	406889	46.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.54%	
Target Compounds						
16) Acetone	2.43	43	29288	7.640	ug/l	99
51) 4-Methyl-2-Pentanone	8.64	43	54806	5.133	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120619\
Data File : VX013789.D
Acq On : 06 Dec 2019 11:11
Operator : JC/SP
Sample : K6099-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OF-3

Quant Time: Dec 07 03:28:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

