

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120619\
 Data File : VX013791.D
 Acq On : 06 Dec 2019 11:59
 Operator : JC/SP
 Sample : K6184-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D1-20191202

Quant Time: Dec 07 03:35:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	657914	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1003596	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	890381	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	397336	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	360766	47.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.48%	
35) Dibromofluoromethane	5.49	113	299595	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
50) Toluene-d8	8.71	98	1179652	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
62) 4-Bromofluorobenzene	11.13	95	399010	45.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.54%	
Target Compounds						
16) Acetone	2.43	43	6136	1.627	ug/l #	88
44) Trichloroethene	7.20	130	35782	4.699	ug/l	91
64) Tetrachloroethene	9.34	164	5905	0.889	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120619\
Data File : VX013791.D
Acq On : 06 Dec 2019 11:59
Operator : JC/SP
Sample : K6184-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20191202

Quant Time: Dec 07 03:35:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

