

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120619\
 Data File : VX013796.D
 Acq On : 06 Dec 2019 13:56
 Operator : JC/SP
 Sample : K6184-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE123D3-20191203

Quant Time: Dec 07 03:52:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	639233	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	981009	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	867645	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	376566	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	350880	47.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.58%	
35) Dibromofluoromethane	5.48	113	294799	48.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.22%	
50) Toluene-d8	8.71	98	1154446	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
62) 4-Bromofluorobenzene	11.14	95	383114	44.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.92%	
Target Compounds						
16) Acetone	2.43	43	10611	2.897	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120619\
Data File : VX013796.D
Acq On : 06 Dec 2019 13:56
Operator : JC/SP
Sample : K6184-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE123D3-20191203

Quant Time: Dec 07 03:52:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

