

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120619\
 Data File : VX013801.D
 Acq On : 06 Dec 2019 15:52
 Operator : JC/SP
 Sample : K6184-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE109D2-20191204

Quant Time: Dec 07 04:06:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	632308	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	971092	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	858630	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	373105	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	346162	47.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.32%	
35) Dibromofluoromethane	5.49	113	288806	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.22%	
50) Toluene-d8	8.71	98	1135913	48.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.12%	
62) 4-Bromofluorobenzene	11.13	95	378032	44.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.64%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	7829	1.299	ug/l	97
16) Acetone	2.44	43	10348	2.856	ug/l #	83
44) Trichloroethene	7.21	130	273225	37.081	ug/l	95
64) Tetrachloroethene	9.34	164	3202	0.500	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120619\
Data File : VX013801.D
Acq On : 06 Dec 2019 15:52
Operator : JC/SP
Sample : K6184-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE109D2-20191204

Quant Time: Dec 07 04:06:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

