

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120619\
 Data File : VX013806.D
 Acq On : 06 Dec 2019 17:49
 Operator : JC/SP
 Sample : K6184-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D1-20191203

Quant Time: Dec 07 04:31:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	638393	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	973798	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	870783	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	374211	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	352099	47.51	ug/l	0.00
Spiked Amount			Recovery	=	95.02%	
35) Dibromofluoromethane	5.49	113	290695	48.29	ug/l	0.00
Spiked Amount			Recovery	=	96.58%	
50) Toluene-d8	8.71	98	1134707	48.38	ug/l	0.00
Spiked Amount			Recovery	=	96.76%	
62) 4-Bromofluorobenzene	11.14	95	374326	44.26	ug/l	0.00
Spiked Amount			Recovery	=	88.52%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.37	101	14895	2.447	ug/l	95
16) Acetone	2.43	43	9208	2.517	ug/l	89
27) cis-1,2-Dichloroethene	4.59	96	10108	1.330	ug/l	87
38) Carbon Tetrachloride	5.78	117	2531	0.311	ug/l #	58
44) Trichloroethene	7.21	130	2729088	369.353	ug/l	97
64) Tetrachloroethene	9.33	164	12321	1.897	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120619\
Data File : VX013806.D
Acq On : 06 Dec 2019 17:49
Operator : JC/SP
Sample : K6184-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D1-20191203

Quant Time: Dec 07 04:31:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

