

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX120619\
 Data File : VX013816.D
 Acq On : 06 Dec 2019 22:28
 Operator : JC/SP
 Sample : K6185-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01-20191205

Quant Time: Dec 07 04:56:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	607949	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	937269	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	814678	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	344406	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	334526	47.40	ug/l	0.00
Spiked Amount						
			Recovery	=	94.80%	
35) Dibromofluoromethane	5.49	113	276422	47.71	ug/l	0.00
Spiked Amount						
			Recovery	=	95.42%	
50) Toluene-d8	8.71	98	1094601	48.49	ug/l	0.00
Spiked Amount						
			Recovery	=	96.98%	
62) 4-Bromofluorobenzene	11.14	95	354317	43.52	ug/l	0.00
Spiked Amount						
			Recovery	=	87.04%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120619\
Data File : VX013816.D
Acq On : 06 Dec 2019 22:28
Operator : JC/SP
Sample : K6185-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-20191205

Quant Time: Dec 07 04:56:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

