

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120619\
 Data File : VX013817.D
 Acq On : 06 Dec 2019 22:51
 Operator : JC/SP
 Sample : K6185-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB01-20191205

Quant Time: Dec 07 04:59:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	630285	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	964526	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	832898	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	354533	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	341788	46.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.44%	
35) Dibromofluoromethane	5.49	113	282933	47.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.90%	
50) Toluene-d8	8.71	98	1119580	48.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.38%	
62) 4-Bromofluorobenzene	11.14	95	368436	43.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.96%	
Target Compounds						
20) Methylene Chloride	2.85	84	7152	1.031	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120619\
Data File : VX013817.D
Acq On : 06 Dec 2019 22:51
Operator : JC/SP
Sample : K6185-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB01-20191205

Quant Time: Dec 07 04:59:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

