

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120619\
 Data File : VX013827.D
 Acq On : 07 Dec 2019 02:44
 Operator : JC/SP
 Sample : K6185-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D3-20191205

Quant Time: Dec 07 05:48:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	605881	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	927181	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	830012	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	361402	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	339023	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
35) Dibromofluoromethane	5.49	113	277777	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
50) Toluene-d8	8.71	98	1103469	49.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.82%	
62) 4-Bromofluorobenzene	11.13	95	364533	45.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.54%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	306808	53.113	ug/l	98
12) 1,1-Dichloroethene	2.37	96	6327	1.097	ug/l	87
16) Acetone	2.43	43	9932	2.860	ug/l	98
27) cis-1,2-Dichloroethene	4.59	96	14685	2.036	ug/l	87
44) Trichloroethene	7.21	130	1330139	189.071	ug/l	98
64) Tetrachloroethene	9.33	164	35868	5.795	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120619\
Data File : VX013827.D
Acq On : 07 Dec 2019 02:44
Operator : JC/SP
Sample : K6185-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D3-20191205

Quant Time: Dec 07 05:48:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

