

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120619\
 Data File : VX013830.D
 Acq On : 07 Dec 2019 03:54
 Operator : JC/SP
 Sample : K6185-01DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D1-20191204DL

Quant Time: Dec 07 05:54:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	613742	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	933856	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	834753	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	365120	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	341258	47.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.80%	
35) Dibromofluoromethane	5.49	113	278035	48.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.32%	
50) Toluene-d8	8.71	98	1106797	49.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.42%	
62) 4-Bromofluorobenzene	11.14	95	371698	45.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.66%	
Target Compounds						
44) Trichloroethene	7.21	130	223759	31.579	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120619\
Data File : VX013830.D
Acq On : 07 Dec 2019 03:54
Operator : JC/SP
Sample : K6185-01DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D1-20191204DL

Quant Time: Dec 07 05:54:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

