

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX120619\
 Data File : VX013810.D
 Acq On : 06 Dec 2019 19:22
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 07 04:40:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.557	0.511	8.3	74	0.00
3 P	Chloromethane	0.624	0.589	5.6	78	0.00
4 C	Vinyl Chloride	0.703	0.692	1.6#	79	0.00
5 T	Bromomethane	0.472	0.447	5.3	84	0.00
6 T	Chloroethane	0.400	0.382	4.5	82	0.00
7 T	Trichlorofluoromethane	0.780	0.738	5.4	82	0.00
8 T	Diethyl Ether	0.353	0.357	-1.1	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.477	0.452	5.2	82	0.00
10 T	Methyl Iodide	0.577	0.581	-0.7	78	0.00
11 T	Tert butyl alcohol	0.143	0.138	3.5	85	0.00
12 CM	1,1-Dichloroethene	0.476	0.459	3.6#	81	0.00
13 T	Acrolein	0.088	0.093	-5.7	80	0.00
14 T	Allyl chloride	0.864	0.839	2.9	81	0.00
15 T	Acrylonitrile	0.296	0.299	-1.0	84	0.00
16 T	Acetone	0.287	0.292	-1.7	85	0.00
17 T	Carbon Disulfide	1.356	1.187	12.5	75	0.00
18 T	Methyl Acetate	0.798	0.869	-8.9	88	0.00
19 T	Methyl tert-butyl Ether	1.576	1.623	-3.0	86	0.00
20 T	Methylene Chloride	0.551	0.548	0.5	87	0.00
21 T	trans-1,2-Dichloroethene	0.518	0.491	5.2	81	0.00
22 T	Diisopropyl ether	1.663	1.731	-4.1	86	0.00
23 T	Vinyl Acetate	1.383	1.350	2.4	80	0.00
24 P	1,1-Dichloroethane	0.909	0.935	-2.9	86	0.00
25 T	2-Butanone	0.426	0.436	-2.3	83	0.00
26 T	2,2-Dichloropropane	0.752	0.658	12.5	74	0.00
27 T	cis-1,2-Dichloroethene	0.595	0.592	0.5	85	0.00
28 T	Bromochloromethane	0.320	0.377	-17.8	85	0.00
29 T	Tetrahydrofuran	0.263	0.270	-2.7	83	0.00
30 C	Chloroform	0.922	0.901	2.3#	84	0.00
31 T	Cyclohexane	0.857	0.785	8.4	78	0.00
32 T	1,1,1-Trichloroethane	0.770	0.762	1.0	83	0.00
33 S	1,2-Dichloroethane-d4	0.580	0.550	5.2	82	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
35 S	Dibromofluoromethane	0.309	0.300	2.9	85	0.00
36 T	1,1-Dichloropropene	0.454	0.427	5.9	80	0.00
37 T	Ethyl Acetate	0.509	0.523	-2.8	83	0.00
38 T	Carbon Tetrachloride	0.419	0.412	1.7	82	0.00
39 T	Methylcyclohexane	0.561	0.507	9.6	77	0.00
40 TM	Benzene	1.382	1.352	2.2	84	0.00
41 T	Methacrylonitrile	0.279	0.289	-3.6	83	0.00
42 TM	1,2-Dichloroethane	0.471	0.477	-1.3	86	0.00
43 T	Isopropyl Acetate	0.838	0.852	-1.7	85	0.00
44 TM	Trichloroethene	0.379	0.358	5.5	81	0.00
45 C	1,2-Dichloropropane	0.346	0.357	-3.2#	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX120619\
 Data File : VX013810.D
 Acq On : 06 Dec 2019 19:22
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 07 04:40:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.235	0.229	2.6	84	0.00
47 T	Bromodichloromethane	0.444	0.459	-3.4	86	0.00
48 T	Methyl methacrylate	0.407	0.423	-3.9	84	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	80	0.00
50 S	Toluene-d8	1.204	1.117	7.2	81	0.00
51 T	4-Methyl-2-Pentanone	0.527	0.546	-3.6	86	0.00
52 CM	Toluene	0.859	0.842	2.0#	83	0.00
53 T	t-1,3-Dichloropropene	0.505	0.500	1.0	81	0.00
54 T	cis-1,3-Dichloropropene	0.552	0.560	-1.4	83	0.00
55 T	1,1,2-Trichloroethane	0.346	0.352	-1.7	86	0.00
56 T	Ethyl methacrylate	0.556	0.580	-4.3	86	0.00
57 T	1,3-Dichloropropane	0.597	0.594	0.5	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.218	0.214	1.8	82	0.00
59 T	2-Hexanone	0.411	0.423	-2.9	85	0.00
60 T	Dibromochloromethane	0.359	0.372	-3.6	85	0.00
61 T	1,2-Dibromoethane	0.370	0.368	0.5	85	0.00
62 S	4-Bromofluorobenzene	0.434	0.411	5.3	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	81	0.00
64 T	Tetrachloroethene	0.373	0.424	-13.7	99	0.00
65 PM	Chlorobenzene	1.038	1.007	3.0	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.383	-2.7	85	0.00
67 C	Ethyl Benzene	1.784	1.784	0.0	82	0.00
68 T	m/p-Xylenes	0.683	0.676	1.0	82	0.00
69 T	o-Xylene	0.666	0.660	0.9	82	0.00
70 T	Styrene	1.120	1.141	-1.9	83	0.00
71 P	Bromoform	0.308	0.326	-5.8	85	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	3.507	3.543	-1.0	83	0.00
74 T	N-amyl acetate	1.607	1.718	-6.9	84	0.00
75 P	1,1,2,2-Tetrachloroethane	1.216	1.271	-4.5	85	0.00
76 T	1,2,3-Trichloropropane	1.125	0.985	12.4	70	0.00
77 T	Bromobenzene	0.948	0.926	2.3	82	0.00
78 T	n-propylbenzene	3.915	3.966	-1.3	82	0.00
79 T	2-Chlorotoluene	2.349	2.403	-2.3	85	0.00
80 T	1,3,5-Trimethylbenzene	2.875	2.945	-2.4	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.411	0.406	1.2	79	0.00
82 T	4-Chlorotoluene	2.720	2.746	-1.0	84	0.00
83 T	tert-Butylbenzene	2.921	2.864	2.0	80	0.00
84 T	1,2,4-Trimethylbenzene	2.917	2.958	-1.4	83	0.00
85 T	sec-Butylbenzene	3.337	3.415	-2.3	83	0.00
86 T	p-Isopropyltoluene	3.104	3.133	-0.9	82	0.00
87 T	1,3-Dichlorobenzene	1.652	1.623	1.8	82	0.00
88 T	1,4-Dichlorobenzene	1.682	1.628	3.2	83	0.00
89 T	n-Butylbenzene	2.648	2.669	-0.8	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX120619\
Data File : VX013810.D
Acq On : 06 Dec 2019 19:22
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 07 04:40:59 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.561	0.563	-0.4	83	0.00
91 T	1,2-Dichlorobenzene	1.632	1.650	-1.1	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.265	0.277	-4.5	83	0.00
93 T	1,2,4-Trichlorobenzene	1.137	1.108	2.6	80	0.00
94 T	Hexachlorobutadiene	0.546	0.511	6.4	82	0.00
95 T	Naphthalene	3.387	3.512	-3.7	81	0.00
96 T	1,2,3-Trichlorobenzene	1.126	1.109	1.5	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5