


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX120619\
Data File : VX013785.D
Acq On : 06 Dec 2019 08:19
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Title : SW846 8260
Last Update : Tue Nov 26 15:53:00 2019



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1671

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	22584	PASS
75	95	30	60	49.7	58372	PASS
95	95	100	100	100.0	117400	PASS
96	95	5	9	7.0	8256	PASS
173	174	0.00	2	1.0	1015	PASS
174	95	50	100	85.4	100289	PASS
175	174	5	9	7.3	7290	PASS
176	174	95	101	98.7	99004	PASS
177	176	5	9	6.6	6566	PASS