

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
 Data File : VX025538.D
 Acq On : 06 Dec 2021 10:51
 Operator : JC/MD
 Sample : VX1206MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1206MBL01

Quant Time: Dec 07 05:55:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	106533	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.757	114	174244	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	156586	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68791	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60407	53.794	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.580%
35) Dibromofluoromethane	5.385	113	53552	50.091	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.180%
50) Toluene-d8	8.647	98	202904	50.478	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.960%
62) 4-Bromofluorobenzene	11.085	95	70659	46.465	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
Data File : VX025538.D
Acq On : 06 Dec 2021 10:51
Operator : JC/MD
Sample : VX1206MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1206MBL01

Quant Time: Dec 07 05:55:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

