

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
 Data File : VX025542.D
 Acq On : 06 Dec 2021 12:34
 Operator : JC/MD
 Sample : M4924-06DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D2-20211202DL

Quant Time: Dec 07 05:57:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	105550	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	173966	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68083	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60891	54.730	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.460%
35) Dibromofluoromethane	5.391	113	53315	49.949	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.900%
50) Toluene-d8	8.653	98	200106	49.861	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.720%
62) 4-Bromofluorobenzene	11.085	95	68418	45.063	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.120%
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	4165	4.240	ug/l	99
12) 1,1-Dichloroethene	2.319	96	3100	3.169	ug/l	90
44) Trichloroethene	7.129	130	65566	49.089	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
Data File : VX025542.D
Acq On : 06 Dec 2021 12:34
Operator : JC/MD
Sample : M4924-06DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D2-20211202DL

Quant Time: Dec 07 05:57:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

