

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
 Data File : VX025543.D
 Acq On : 06 Dec 2021 12:57
 Operator : JC/MD
 Sample : M4923-02DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D1-20211130DL

Quant Time: Dec 07 05:58:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) Pentafluorobenzene		5.556	168	100353	50.000	ug/l	0.00
34) 1,4-Difluorobenzene		6.763	114	166898	50.000	ug/l	0.00
63) Chlorobenzene-d5		10.055	117	157749	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4		12.024	152	72431	50.000	ug/l	0.00
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4		5.958	65	58271	55.088	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.180%	
35) Dibromofluoromethane		5.391	113	50951	49.755	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.520%	
50) Toluene-d8		8.653	98	193598	50.282	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.560%	
62) 4-Bromofluorobenzene		11.085	95	73346	50.355	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.700%	
Target Compounds							
							Qvalue
9) 1,1,2-Trichlorotrifluo...		2.319	101	2247	2.406	ug/l	98
44) Trichloroethene		7.129	130	45118	35.210	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
Data File : VX025543.D
Acq On : 06 Dec 2021 12:57
Operator : JC/MD
Sample : M4923-02DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D1-20211130DL

Quant Time: Dec 07 05:58:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

