

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
 Data File : VX025554.D
 Acq On : 06 Dec 2021 17:16
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Dec 07 06:02:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	95530	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	156580	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	139660	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61527	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55216	54.835	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.660%
35) Dibromofluoromethane	5.391	113	48057	50.022	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.040%
50) Toluene-d8	8.653	98	178830	49.507	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.020%
62) 4-Bromofluorobenzene	11.085	95	61728	45.171	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
Data File : VX025554.D
Acq On : 06 Dec 2021 17:16
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Dec 07 06:02:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

