

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
 Data File : VX025555.D
 Acq On : 06 Dec 2021 17:40
 Operator : JC/MD
 Sample : PB141182ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141182ZHE#05

Quant Time: Dec 07 06:03:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	98161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	160388	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	147386	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65942	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	56799	54.895	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.800%
35) Dibromofluoromethane	5.391	113	48959	49.751	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.500%
50) Toluene-d8	8.653	98	187309	50.624	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.240%
62) 4-Bromofluorobenzene	11.085	95	65520	46.808	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.620%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	10498	18.260	ug/l	94
20) Methylene Chloride	2.794	84	3236	2.994	ug/l	94
43) Isopropyl Acetate	6.349	43	8351	3.152	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
Data File : VX025555.D
Acq On : 06 Dec 2021 17:40
Operator : JC/MD
Sample : PB141182ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141182ZHE#05

Quant Time: Dec 07 06:03:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

