

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
 Data File : VX025557.D
 Acq On : 06 Dec 2021 18:27
 Operator : JC/MD
 Sample : PB141182ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141182ZHE#07

Quant Time: Dec 07 06:04:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	93479	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	154110	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	138955	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58553	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55644	56.472	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.940%
35) Dibromofluoromethane	5.391	113	48499	51.291	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.580%
50) Toluene-d8	8.653	98	178283	50.147	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.300%
62) 4-Bromofluorobenzene	11.085	95	59313	44.100	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.200%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	10726	19.590	ug/l	93
20) Methylene Chloride	2.788	84	3436	3.338	ug/l	97
25) 2-Butanone	4.574	43	3197	4.039	ug/l	90
43) Isopropyl Acetate	6.348	43	8613	3.383	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
Data File : VX025557.D
Acq On : 06 Dec 2021 18:27
Operator : JC/MD
Sample : PB141182ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141182ZHE#07

Quant Time: Dec 07 06:04:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

