

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
 Data File : VX025559.D
 Acq On : 06 Dec 2021 19:15
 Operator : JC/MD
 Sample : PB141182ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141182ZHE#09

Quant Time: Dec 07 06:05:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	95839	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	158440	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	145464	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64095	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56285	55.716	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 111.440%			
35) Dibromofluoromethane	5.391	113	49124	50.532	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.060%			
50) Toluene-d8	8.653	98	183480	50.198	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.400%			
62) 4-Bromofluorobenzene	11.085	95	64046	46.317	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 92.640%			
Target Compounds						
16) Acetone	2.386	43	10579	18.846	ug/l	97
20) Methylene Chloride	2.794	84	3355	3.179	ug/l	95
25) 2-Butanone	4.574	43	3165	3.900	ug/l	98
43) Isopropyl Acetate	6.348	43	8836	3.376	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
Data File : VX025559.D
Acq On : 06 Dec 2021 19:15
Operator : JC/MD
Sample : PB141182ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141182ZHE#09

Quant Time: Dec 07 06:05:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

