

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
 Data File : VX025560.D
 Acq On : 06 Dec 2021 19:38
 Operator : JC/MD
 Sample : PB141182ZHE#10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141182ZHE#10

Quant Time: Dec 07 06:05:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	92916	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	153177	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141303	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60338	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55072	56.230	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.460%
35) Dibromofluoromethane	5.391	113	47591	50.637	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.280%
50) Toluene-d8	8.653	98	179809	50.884	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.760%
62) 4-Bromofluorobenzene	11.085	95	61401	45.930	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.860%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	10687	19.638	ug/l	95
20) Methylene Chloride	2.794	84	3416	3.339	ug/l	94
25) 2-Butanone	4.580	43	3097	3.936	ug/l	91
43) Isopropyl Acetate	6.354	43	8659	3.422	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
Data File : VX025560.D
Acq On : 06 Dec 2021 19:38
Operator : JC/MD
Sample : PB141182ZHE#10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141182ZHE#10

Quant Time: Dec 07 06:05:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

