

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX120621\
 Data File : VX025537.D
 Acq On : 06 Dec 2021 10:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 07 05:54:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	68	0.00
2 T	Dichlorodifluoromethane	50.000	55.225	-10.5	66	0.00
3 P	Chloromethane	50.000	44.939	10.1	59	0.00
4 C	Vinyl Chloride	50.000	46.243	7.5#	59	0.00
5 T	Bromomethane	50.000	59.798	-19.6	80	0.00
6 T	Chloroethane	50.000	44.856	10.3	63	0.00
7 T	Trichlorofluoromethane	50.000	51.249	-2.5	67	0.00
8 T	Diethyl Ether	50.000	48.431	3.1	66	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	56.344	-12.7	76	0.00
10 T	Methyl Iodide	50.000	39.870	20.3	49	0.00
11 T	Tert butyl alcohol	250.000	220.075	12.0	58	0.00
12 CM	1,1-Dichloroethene	50.000	52.262	-4.5#	70	0.00
13 T	Acrolein	250.000	254.725	-1.9	66	0.00
14 T	Allyl chloride	50.000	48.955	2.1	69	0.00
15 T	Acrylonitrile	250.000	254.041	-1.6	68	0.00
16 T	Acetone	250.000	286.501	-14.6	77	0.00
17 T	Carbon Disulfide	50.000	42.363	15.3	59	0.00
18 T	Methyl Acetate	50.000	49.174	1.7	68	0.00
19 T	Methyl tert-butyl Ether	50.000	53.033	-6.1	73	0.00
20 T	Methylene Chloride	50.000	49.746	0.5	72	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.961	-1.9	70	0.00
22 T	Diisopropyl ether	50.000	52.765	-5.5	72	0.00
23 T	Vinyl Acetate	250.000	261.410	-4.6	71	0.00
24 P	1,1-Dichloroethane	50.000	53.262	-6.5	73	0.00
25 T	2-Butanone	250.000	259.464	-3.8	69	0.00
26 T	2,2-Dichloropropane	50.000	52.361	-4.7	71	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.934	-5.9	72	0.00
28 T	Bromochloromethane	50.000	50.774	-1.5	68	0.00
29 T	Tetrahydrofuran	250.000	242.435	3.0	66	0.00
30 C	Chloroform	50.000	55.201	-10.4#	75	0.00
31 T	Cyclohexane	50.000	50.356	-0.7	69	0.00
32 T	1,1,1-Trichloroethane	50.000	53.445	-6.9	73	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.622	0.8	68	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	67	0.00
35 S	Dibromofluoromethane	50.000	49.443	1.1	67	0.00
36 T	1,1-Dichloropropene	50.000	53.119	-6.2	72	0.00
37 T	Ethyl Acetate	50.000	50.072	-0.1	68	0.00
38 T	Carbon Tetrachloride	50.000	54.354	-8.7	73	0.00
39 T	Methylcyclohexane	50.000	52.418	-4.8	70	0.00
40 TM	Benzene	50.000	52.649	-5.3	72	0.00
41 T	Methacrylonitrile	50.000	49.761	0.5	70	0.00
42 TM	1,2-Dichloroethane	50.000	56.545	-13.1	76	0.00
43 T	Isopropyl Acetate	50.000	50.486	-1.0	68	0.00
44 TM	Trichloroethene	50.000	53.879	-7.8	72	0.00
45 C	1,2-Dichloropropane	50.000	53.810	-7.6#	72	0.00
46 T	Dibromomethane	50.000	55.057	-10.1	74	0.00
47 T	Bromodichloromethane	50.000	54.459	-8.9	73	0.00
48 T	Methyl methacrylate	50.000	50.205	-0.4	69	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
 Data File : VX025537.D
 Acq On : 06 Dec 2021 10:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 07 05:54:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	936.106	6.4	61	0.00
50 S	Toluene-d8	50.000	47.291	5.4	63	0.00
51 T	4-Methyl-2-Pentanone	250.000	254.821	-1.9	68	0.00
52 CM	Toluene	50.000	52.692	-5.4#	71	0.00
53 T	t-1,3-Dichloropropene	50.000	52.637	-5.3	71	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.828	-5.7	71	0.00
55 T	1,1,2-Trichloroethane	50.000	55.599	-11.2	74	0.00
56 T	Ethyl methacrylate	50.000	51.194	-2.4	69	0.00
57 T	1,3-Dichloropropane	50.000	54.762	-9.5	74	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	271.160	-8.5	74	0.00
59 T	2-Hexanone	250.000	256.689	-2.7	68	0.00
60 T	Dibromochloromethane	50.000	54.395	-8.8	73	0.00
61 T	1,2-Dibromoethane	50.000	53.993	-8.0	73	0.00
62 S	4-Bromofluorobenzene	50.000	48.936	2.1	67	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	67	0.00
64 T	Tetrachloroethene	50.000	55.460	-10.9	75	0.00
65 PM	Chlorobenzene	50.000	54.651	-9.3	74	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	54.100	-8.2	73	0.00
67 C	Ethyl Benzene	50.000	54.694	-9.4#	73	0.00
68 T	m/p-Xylenes	100.000	107.331	-7.3	73	0.00
69 T	o-Xylene	50.000	54.000	-8.0	73	0.00
70 T	Styrene	50.000	54.990	-10.0	73	0.00
71 P	Bromoform	50.000	50.489	-1.0	67	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	68	0.00
73 T	Isopropylbenzene	50.000	52.867	-5.7	74	0.00
74 T	N-ethyl acetate	50.000	49.031	1.9	68	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.343	-2.7	72	0.00
76 T	1,2,3-Trichloropropane	50.000	51.968	-3.9	71	0.00
77 T	Bromobenzene	50.000	52.614	-5.2	73	0.00
78 T	n-propylbenzene	50.000	54.892	-9.8	75	0.00
79 T	2-Chlorotoluene	50.000	52.880	-5.8	75	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.275	-6.5	75	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	46.206	7.6	63	0.00
82 T	4-Chlorotoluene	50.000	53.639	-7.3	75	0.00
83 T	tert-Butylbenzene	50.000	52.636	-5.3	74	0.00
84 T	1,2,4-Trimethylbenzene	50.000	54.236	-8.5	75	0.00
85 T	sec-Butylbenzene	50.000	55.953	-11.9	77	0.00
86 T	p-Isopropyltoluene	50.000	55.087	-10.2	76	0.00
87 T	1,3-Dichlorobenzene	50.000	54.546	-9.1	75	0.00
88 T	1,4-Dichlorobenzene	50.000	54.321	-8.6	76	0.00
89 T	n-Butylbenzene	50.000	57.317	-14.6	79	0.00
90 T	Hexachloroethane	50.000	49.771	0.5	69	0.00
91 T	1,2-Dichlorobenzene	50.000	54.229	-8.5	75	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.441	-0.9	66	0.00
93 T	1,2,4-Trichlorobenzene	50.000	61.648	-23.3	77	0.00
94 T	Hexachlorobutadiene	50.000	56.533	-13.1	80	0.00
95 T	Naphthalene	50.000	55.054	-10.1	72	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
Data File : VX025537.D
Acq On : 06 Dec 2021 10:14
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 07 05:54:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	61.847	-23.7	79	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6