

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX120621\
 Data File : VX025564.D
 Acq On : 06 Dec 2021 21:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 07 06:07:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	56	0.00
2 T	Dichlorodifluoromethane	50.000	51.072	-2.1	50	0.00
3 P	Chloromethane	50.000	46.079	7.8	50	0.00
4 C	Vinyl Chloride	50.000	45.348	9.3#	48	0.00
5 T	Bromomethane	50.000	56.728	-13.5	62	0.00
6 T	Chloroethane	50.000	47.115	5.8	55	0.00
7 T	Trichlorofluoromethane	50.000	49.547	0.9	53	0.00
8 T	Diethyl Ether	50.000	53.361	-6.7	60	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	52.122	-4.2	58	0.00
10 T	Methyl Iodide	50.000	49.833	0.3	51	0.00
11 T	Tert butyl alcohol	250.000	279.923	-12.0	60	0.01
12 CM	1,1-Dichloroethene	50.000	51.129	-2.3#	56	0.00
13 T	Acrolein	250.000	304.821	-21.9	65	0.00
14 T	Allyl chloride	50.000	50.392	-0.8	58	0.00
15 T	Acrylonitrile	250.000	303.951	-21.6	67	0.00
16 T	Acetone	250.000	247.857	0.9	55	0.00
17 T	Carbon Disulfide	50.000	38.079	23.8	44	0.00
18 T	Methyl Acetate	50.000	59.064	-18.1	68	0.00
19 T	Methyl tert-butyl Ether	50.000	59.308	-18.6	67	0.00
20 T	Methylene Chloride	50.000	54.438	-8.9	65	0.00
21 T	trans-1,2-Dichloroethene	50.000	52.080	-4.2	59	0.00
22 T	Diisopropyl ether	50.000	57.735	-15.5	65	0.00
23 T	Vinyl Acetate	250.000	288.608	-15.4	64	0.00
24 P	1,1-Dichloroethane	50.000	57.556	-15.1	65	0.00
25 T	2-Butanone	250.000	284.590	-13.8	63	0.01
26 T	2,2-Dichloropropane	50.000	43.141	13.7	48	0.00
27 T	cis-1,2-Dichloroethene	50.000	57.733	-15.5	65	0.00
28 T	Bromochloromethane	50.000	57.161	-14.3	63	0.00
29 T	Tetrahydrofuran	250.000	294.421	-17.8	66	0.00
30 C	Chloroform	50.000	59.782	-19.6#	67	0.00
31 T	Cyclohexane	50.000	46.758	6.5	53	0.00
32 T	1,1,1-Trichloroethane	50.000	54.745	-9.5	62	0.00
33 S	1,2-Dichloroethane-d4	50.000	57.959	-15.9	65	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	59	0.00
35 S	Dibromofluoromethane	50.000	52.347	-4.7	63	0.00
36 T	1,1-Dichloropropene	50.000	49.191	1.6	59	0.00
37 T	Ethyl Acetate	50.000	55.692	-11.4	67	0.00
38 T	Carbon Tetrachloride	50.000	49.508	1.0	59	0.00
39 T	Methylcyclohexane	50.000	43.363	13.3	51	0.00
40 TM	Benzene	50.000	52.434	-4.9	63	0.00
41 T	Methacrylonitrile	50.000	55.287	-10.6	69	0.00
42 TM	1,2-Dichloroethane	50.000	58.708	-17.4	70	0.00
43 T	Isopropyl Acetate	50.000	54.360	-8.7	65	0.00
44 TM	Trichloroethene	50.000	50.764	-1.5	60	0.00
45 C	1,2-Dichloropropane	50.000	54.516	-9.0#	64	0.00
46 T	Dibromomethane	50.000	56.320	-12.6	67	0.00
47 T	Bromodichloromethane	50.000	53.745	-7.5	64	0.00
48 T	Methyl methacrylate	50.000	53.736	-7.5	65	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX120621\
 Data File : VX025564.D
 Acq On : 06 Dec 2021 21:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 07 06:07:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1148.014	-14.8	66	0.00
50 S	Toluene-d8	50.000	49.936	0.1	59	0.00
51 T	4-Methyl-2-Pentanone	250.000	284.451	-13.8	67	0.00
52 CM	Toluene	50.000	52.452	-4.9#	62	0.00
53 T	t-1,3-Dichloropropene	50.000	49.823	0.4	59	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.081	-0.2	59	0.00
55 T	1,1,2-Trichloroethane	50.000	57.756	-15.5	68	0.00
56 T	Ethyl methacrylate	50.000	54.447	-8.9	65	0.00
57 T	1,3-Dichloropropane	50.000	56.245	-12.5	67	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	285.046	-14.0	69	0.00
59 T	2-Hexanone	250.000	274.082	-9.6	64	0.00
60 T	Dibromochloromethane	50.000	51.706	-3.4	61	0.00
61 T	1,2-Dibromoethane	50.000	55.186	-10.4	66	0.00
62 S	4-Bromofluorobenzene	50.000	51.136	-2.3	62	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	60	0.00
64 T	Tetrachloroethene	50.000	44.109	11.8	53	0.00
65 PM	Chlorobenzene	50.000	52.696	-5.4	64	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	53.038	-6.1	64	0.00
67 C	Ethyl Benzene	50.000	51.738	-3.5#	62	0.00
68 T	m/p-Xylenes	100.000	101.658	-1.7	61	0.00
69 T	o-Xylene	50.000	52.379	-4.8	63	0.00
70 T	Styrene	50.000	53.303	-6.6	63	0.00
71 P	Bromoform	50.000	48.479	3.0	57	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	59	0.00
73 T	Isopropylbenzene	50.000	51.390	-2.8	62	0.00
74 T	N-ethyl acetate	50.000	53.848	-7.7	65	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	56.755	-13.5	69	0.00
76 T	1,2,3-Trichloropropane	50.000	53.969	-7.9	64	0.00
77 T	Bromobenzene	50.000	54.237	-8.5	65	0.00
78 T	n-propylbenzene	50.000	51.294	-2.6	60	0.00
79 T	2-Chlorotoluene	50.000	52.115	-4.2	64	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.112	-2.2	62	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	40.450	19.1	47	0.00
82 T	4-Chlorotoluene	50.000	51.616	-3.2	62	0.00
83 T	tert-Butylbenzene	50.000	51.941	-3.9	63	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.108	-6.2	63	0.00
85 T	sec-Butylbenzene	50.000	50.860	-1.7	60	0.00
86 T	p-Isopropyltoluene	50.000	50.066	-0.1	59	0.00
87 T	1,3-Dichlorobenzene	50.000	52.767	-5.5	62	0.00
88 T	1,4-Dichlorobenzene	50.000	51.751	-3.5	63	0.00
89 T	n-Butylbenzene	50.000	48.929	2.1	58	0.00
90 T	Hexachloroethane	50.000	43.808	12.4	52	0.00
91 T	1,2-Dichlorobenzene	50.000	52.875	-5.8	63	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	56.968	-13.9	65	0.00
93 T	1,2,4-Trichlorobenzene	50.000	57.056	-14.1	62	0.00
94 T	Hexachlorobutadiene	50.000	45.594	8.8	55	0.00
95 T	Naphthalene	50.000	58.403	-16.8	66	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120621\
Data File : VX025564.D
Acq On : 06 Dec 2021 21:13
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 07 06:07:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	58.954	-17.9	65	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6