

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120622\
 Data File : VX033255.D
 Acq On : 06 Dec 2022 11:46
 Operator : JC/MD
 Sample : VX1206MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1206MBL01

Quant Time: Dec 07 01:30:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	137067	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	215158	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186202	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85986	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	45186	32.650	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	65.300%#
35) Dibromofluoromethane	5.385	113	49857	41.103	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	82.200%
50) Toluene-d8	8.647	98	103420	27.360	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	54.720%#
62) 4-Bromofluorobenzene	11.079	95	77490	48.633	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120622\
Data File : VX033255.D
Acq On : 06 Dec 2022 11:46
Operator : JC/MD
Sample : VX1206MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1206MBL01

Quant Time: Dec 07 01:30:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

