

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120622\
 Data File : VX033267.D
 Acq On : 06 Dec 2022 18:32
 Operator : JC/MD
 Sample : N5875-10DL 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-44DL

Quant Time: Dec 07 01:35:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	106648	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	163583	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	168155	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98815	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	99221	98.234	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 196.460%#	
35) Dibromofluoromethane	5.385	113	76353	82.793	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 165.580%#	
50) Toluene-d8	8.647	98	274028	100.955	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 201.920%#	
62) 4-Bromofluorobenzene	11.079	95	93981	77.578	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 155.160%#	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.995	59	20682	125.622	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	33558	10.397	ug/l #	89
39) Methylcyclohexane	7.379	83	1559	0.927	ug/l	99
40) Benzene	6.044	78	4038	0.977	ug/l	95
67) Ethyl Benzene	10.195	91	32793	5.514	ug/l	100
68) m/p-Xylenes	10.305	106	1637	0.703	ug/l	87
73) Isopropylbenzene	10.963	105	4829	0.684	ug/l	100
78) n-propylbenzene	11.305	91	12450	1.542	ug/l	98
95) Naphthalene	13.774	128	11200	1.773	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120622\
Data File : VX033267.D
Acq On : 06 Dec 2022 18:32
Operator : JC/MD
Sample : N5875-10DL 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-44DL

Quant Time: Dec 07 01:35:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

