

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120622\
 Data File : VX033259.D
 Acq On : 06 Dec 2022 15:08
 Operator : JC/MD
 Sample : VX1206WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1206WBSD02

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 12/07/2022
 Supervised By :Mahesh Dadoda 12/07/2022

Quant Time: Dec 07 01:32:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	139683	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	218668	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	197476	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101506	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73214	53.883	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 107.760%			
35) Dibromofluoromethane	5.385	113	65833	53.403	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.800%			
50) Toluene-d8	8.647	98	183710	49.507	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 99.020%			
62) 4-Bromofluorobenzene	11.079	95	85903	53.047	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 106.100%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	17137	16.058	ug/l	96
3) Chloromethane	1.294	50	19160	16.631	ug/l	99
4) Vinyl Chloride	1.374	62	21626	16.859	ug/l	98
5) Bromomethane	1.617	94	15980	16.800	ug/l	99
6) Chloroethane	1.691	64	16667	15.960	ug/l	99
7) Trichlorofluoromethane	1.892	101	45303	18.197	ug/l	98
8) Diethyl Ether	2.130	74	16386	20.906	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.331	101	25181	19.243	ug/l	97
10) Methyl Iodide	2.453	142	23598	15.753	ug/l	99
11) Tert butyl alcohol	3.001	59	21874	101.440	ug/l	100
12) 1,1-Dichloroethene	2.325	96	22935	18.815	ug/l	99
13) Acrolein	2.239	56	29518	119.174	ug/l	98
14) Allyl chloride	2.666	41	37843	19.229	ug/l	96
15) Acrylonitrile	3.068	53	63030	103.342	ug/l	99
16) Acetone	2.392	43	67441	106.158	ug/l	99
17) Carbon Disulfide	2.514	76	53032	15.747	ug/l	99
18) Methyl Acetate	2.709	43	36768	20.908	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	87678	20.740	ug/l	98
20) Methylene Chloride	2.788	84	27810	20.267	ug/l	93
21) trans-1,2-Dichloroethene	3.093	96	25460	18.259	ug/l	96
22) Diisopropyl ether	3.757	45	84694	19.764	ug/l #	86
23) Vinyl Acetate	3.721	43	324307	96.684	ug/l	98
24) 1,1-Dichloroethane	3.611	63	47821	19.502	ug/l	99
25) 2-Butanone	4.562	43	92100	105.136	ug/l	98
26) 2,2-Dichloropropane	4.477	77	37680	19.111	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	31559	19.920	ug/l	99
28) Bromochloromethane	4.903	49	22473	20.857	ug/l	97
29) Tetrahydrofuran	5.013	42	55236	99.370	ug/l	96
30) Chloroform	5.092	83	55688	20.277	ug/l	99
31) Cyclohexane	5.470	56	39730	18.025	ug/l	94
32) 1,1,1-Trichloroethane	5.385	97	48650	19.780	ug/l	99
36) 1,1-Dichloropropene	5.696	75	36984	18.763	ug/l	99
37) Ethyl Acetate	4.714	43	34389	19.997	ug/l	100
38) Carbon Tetrachloride	5.678	117	42424	19.401	ug/l	97
39) Methylcyclohexane	7.379	83	42125	18.730	ug/l	97
40) Benzene	6.037	78	107852	19.515	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120622\
 Data File : VX033259.D
 Acq On : 06 Dec 2022 15:08
 Operator : JC/MD
 Sample : VX1206WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1206WBSD02

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 12/07/2022
 Supervised By :Mahesh Dadoda 12/07/2022

Quant Time: Dec 07 01:32:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	20297	20.976	ug/l	98
42) 1,2-Dichloroethane	6.092	62	44460	19.980	ug/l	98
43) Isopropyl Acetate	6.342	43	57005	20.043	ug/l	99
44) Trichloroethene	7.123	130	30192	19.468	ug/l	99
45) 1,2-Dichloropropane	7.433	63	27826	20.373	ug/l	93
46) Dibromomethane	7.586	93	21599	20.267	ug/l	98
47) Bromodichloromethane	7.824	83	42713	20.742	ug/l	97
48) Methyl methacrylate	7.696	41	29003	19.811	ug/l	96
49) 1,4-Dioxane	7.702	88	12162	451.310	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	184187	104.951	ug/l	98
52) Toluene	8.720	92	71442	19.649	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	42615	20.193	ug/l	96
54) cis-1,3-Dichloropropene	8.366	75	45378	20.441	ug/l	94
55) 1,1,2-Trichloroethane	9.153	97	31349	21.177	ug/l	97
56) Ethyl methacrylate	9.116	69	42872	21.020	ug/l	94
57) 1,3-Dichloropropane	9.305	76	48976	20.494	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	102257	105.579	ug/l	99
59) 2-Hexanone	9.433	43	141493	109.140	ug/l	98
60) Dibromochloromethane	9.518	129	34725	21.788	ug/l	99
61) 1,2-Dibromoethane	9.610	107	32021	20.494	ug/l	98
64) Tetrachloroethene	9.275	164	26926	19.263	ug/l	99
65) Chlorobenzene	10.079	112	80833	20.531	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	31579	21.890	ug/l	97
67) Ethyl Benzene	10.195	91	140754	20.154	ug/l	99
68) m/p-Xylenes	10.299	106	110578	40.433	ug/l	99
69) o-Xylene	10.640	106	54242	20.498	ug/l	100
70) Styrene	10.652	104	90613	20.931	ug/l	98
71) Bromoform	10.799	173	25105	22.698	ug/l #	99
73) Isopropylbenzene	10.963	105	145285	20.030	ug/l	100
74) N-amyl acetate	10.841	43	49059	19.882	ug/l	96
75) 1,1,2,2-Tetrachloroethane	11.213	83	46095	21.055	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	40749m	21.197	ug/l	
77) Bromobenzene	11.195	156	36078	20.224	ug/l	99
78) n-propylbenzene	11.305	91	165419	19.946	ug/l	99
79) 2-Chlorotoluene	11.366	91	99572	19.768	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	122896	20.169	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	12255	20.722	ug/l	98
82) 4-Chlorotoluene	11.457	91	117051	19.830	ug/l	99
83) tert-Butylbenzene	11.713	119	124321	20.496	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	124432	20.480	ug/l	100
85) sec-Butylbenzene	11.890	105	147168	20.281	ug/l	100
86) p-Isopropyltoluene	12.012	119	127180	20.686	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	68440	20.455	ug/l	99
88) 1,4-Dichlorobenzene	12.042	146	69045	20.149	ug/l	99
89) n-Butylbenzene	12.335	91	106174	20.046	ug/l	99
90) Hexachloroethane	12.536	117	21238	21.026	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	67651	20.785	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	9901	22.667	ug/l	98
93) 1,2,4-Trichlorobenzene	13.591	180	41993	20.930	ug/l	98
94) Hexachlorobutadiene	13.725	225	17420	19.680	ug/l	99
95) Naphthalene	13.774	128	138687	21.367	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	42668	21.207	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120622\
Data File : VX033259.D
Acq On : 06 Dec 2022 15:08
Operator : JC/MD
Sample : VX1206WBSD02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022

Quant Time: Dec 07 01:32:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02



Tgt Ion:168 Resp: 13968
Ion Ratio Lower Upper
168 100
99 58.7 46.1 69.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#2
Dichlorodifluoromethane
Concen: 16.058 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Tgt Ion: 85 Resp: 17137
Ion Ratio Lower Upper
85 100
87 33.1 15.6 46.7









#3
Chloromethane
Concen: 16.631 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02

Tgt Ion: 50 Resp: 19160
Ion Ratio Lower Upper
50 100
52 31.2 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#4
Vinyl Chloride
Concen: 16.859 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Tgt Ion: 62 Resp: 21626
Ion Ratio Lower Upper
62 100
64 33.3 25.8 38.6









#5
Bromomethane
Concen: 16.800 ug/l
RT: 1.617 min Scan# 81
Delta R.T. 0.012 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02

Tgt Ion: 94 Resp: 15980
Ion Ratio Lower Upper
94 100
96 95.9 77.6 116.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#6
Chloroethane
Concen: 15.960 ug/l
RT: 1.691 min Scan# 99
Delta R.T. 0.006 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08


Tgt Ion: 64 Resp: 16667
Ion Ratio Lower Upper
64 100
66 32.8 26.0 39.0





#7
Trichlorofluoromethane
Concen: 18.197 ug/l
RT: 1.892 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02



Tgt Ion:101 Resp: 4530
Ion Ratio Lower Upper
101 100
103 63.0 51.4 77.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#8
Diethyl Ether
Concen: 20.906 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Tgt Ion: 74 Resp: 16386
Ion Ratio Lower Upper
74 100
45 87.9 46.6 139.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.243 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02

Tgt Ion:101 Resp: 2518
Ion Ratio Lower Upper
101 100
85 46.3 37.0 55.4
151 84.7 64.4 96.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#10
Methyl Iodide
Concen: 15.753 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Tgt Ion:142 Resp: 23598
Ion Ratio Lower Upper
142 100
127 47.8 38.6 57.8
141 14.7 11.4 17.2





#11

Tert butyl alcohol

Concen: 101.440 ug/l

RT: 3.001 min Scan# 31

Delta R.T. 0.036 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02

Tgt Ion: 59 Resp: 21874

Ion Ratio Lower Upper

59 100





57 10.2 8.2 12.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/07/2022

Supervised By :Mahesh Dadoda 12/07/2022



#12

1,1-Dichloroethene

Concen: 18.815 ug/l

RT: 2.325 min Scan# 203

Delta R.T. 0.006 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08




Tgt Ion: 96 Resp: 22935


Ion Ratio Lower Upper

96 100

61 155.9 125.1 187.7

98 65.2 50.6 76.0





#13

Acrolein

Concen: 119.174 ug/l

RT: 2.239 min Scan# 189

Delta R.T. 0.006 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02

Tgt Ion: 56 Resp: 29518

Ion Ratio Lower Upper

56 100





55 72.4 56.9 85.3

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/07/2022

Supervised By :Mahesh Dadoda 12/07/2022



#14

Allyl chloride

Concen: 19.229 ug/l

RT: 2.666 min Scan# 259

Delta R.T. 0.000 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08




Tgt Ion: 41 Resp: 37843


Ion Ratio Lower Upper

41 100

39 72.7 61.4 92.2


76 34.4 29.2 43.8





#15
Acrylonitrile
Concen: 103.342 ug/l
RT: 3.068 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02




Tgt Ion: 53 Resp: 63030
Ion Ratio Lower Upper
53 100
52 82.6 67.4 101.0
51 38.4 30.3 45.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022




#16
Acetone
Concen: 106.158 ug/l
RT: 2.392 min Scan# 214
Delta R.T. 0.012 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08



Tgt Ion: 43 Resp: 67441
Ion Ratio Lower Upper
43 100
58 30.4 23.8 35.8









#17
Carbon Disulfide
Concen: 15.747 ug/l
RT: 2.514 min Scan# 211
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02

Tgt Ion: 76 Resp: 53032
Ion Ratio Lower Upper
76 100
78 9.0 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#18
Methyl Acetate
Concen: 20.908 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.006 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Tgt Ion: 43 Resp: 36768
Ion Ratio Lower Upper
43 100
74 24.7 19.4 29.0









#19
Methyl tert-butyl Ether
Concen: 20.740 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02

Tgt Ion: 73 Resp: 87678
Ion Ratio Lower Upper
73 100
57 19.9 16.6 25.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#20
Methylene Chloride
Concen: 20.267 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08


Tgt Ion: 84 Resp: 27810
Ion Ratio Lower Upper
84 100
49 116.8 101.0 151.6
51 36.9 31.1 46.7
86 59.8 52.3 78.5





#21
trans-1,2-Dichloroethene
Concen: 18.259 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02




Tgt Ion: 96 Resp: 25460
Ion Ratio Lower Upper
96 100
61 137.2 106.6 160.0
98 60.1 51.0 76.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022




#22
Diisopropyl ether
Concen: 19.764 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Tgt Ion: 45 Resp: 84694
Ion Ratio Lower Upper
45 100
43 88.6 56.4 84.6#
87 31.4 24.4 36.6
59 12.0 9.5 14.3





#23
 Vinyl Acetate
 Concen: 96.684 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX033259.D
 Acq: 06 Dec 2022 15:08




Instrument :
 MSVOA_X
 ClientSampleId :
 VX1206WBSD02




Tgt Ion: 43 Resp: 32430
 Ion Ratio Lower Upper
 43 100
 86 11.4 8.6 12.8

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 12/07/2022
 Supervised By :Mahesh Dadoda 12/07/2022




#24
 1,1-Dichloroethane
 Concen: 19.502 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX033259.D
 Acq: 06 Dec 2022 15:08



Tgt Ion: 63 Resp: 47821
 Ion Ratio Lower Upper
 63 100
 98 6.7 3.5 10.4
 100 3.9 2.3 6.8





#25

2-Butanone

Concen: 105.136 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02

Tgt Ion: 43 Resp: 92100

Ion Ratio Lower Upper

43 100





72 26.8 22.2 33.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/07/2022

Supervised By :Mahesh Dadoda 12/07/2022



#26

2,2-Dichloropropane

Concen: 19.111 ug/l

RT: 4.477 min Scan# 556

Delta R.T. -0.000 min

Lab File: VX033259.D




Acq: 06 Dec 2022 15:08


Tgt Ion: 77 Resp: 37680

Ion Ratio Lower Upper

77 100

97 23.1 11.4 34.1





#27

cis-1,2-Dichloroethene

Concen: 19.920 ug/l

RT: 4.483 min Scan# 51

Delta R.T. -0.006 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02

Tgt Ion: 96 Resp: 3155

Ion Ratio Lower Upper

96 100

61 141.6 0.0 286.0





98 65.5 0.0 129.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/07/2022

Supervised By :Mahesh Dadoda 12/07/2022



#28

Bromochloromethane

Concen: 20.857 ug/l

RT: 4.903 min Scan# 626

Delta R.T. -0.001 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08




Tgt Ion: 49 Resp: 22473


Ion Ratio Lower Upper

49 100

129 2.5 0.0 3.8

130 87.3 67.5 101.3





#29

Tetrahydrofuran

Concen: 99.370 ug/l

RT: 5.013 min Scan# 644

Delta R.T. 0.012 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02

Tgt Ion: 42 Resp: 55230





Ion	Ratio	Lower	Upper
42	100		
72	50.1	37.6	56.4
71	45.9	34.7	52.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/07/2022

Supervised By :Mahesh Dadoda 12/07/2022



#30

Chloroform

Concen: 20.277 ug/l

RT: 5.092 min Scan# 657




Delta R.T. -0.007 min


Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Tgt Ion: 83 Resp: 55688


Ion	Ratio	Lower	Upper
83	100		
85	63.3	51.1	76.7





#31
Cyclohexane
Concen: 18.025 ug/l
RT: 5.470 min Scan# 719
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02




Tgt Ion: 56 Resp: 39730
Ion Ratio Lower Upper
56 100
69 30.7 24.1 36.1
84 90.2 66.9 100.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022




#32
1,1,1-Trichloroethane
Concen: 19.780 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08



Tgt Ion: 97 Resp: 48650
Ion Ratio Lower Upper
97 100
99 63.6 51.4 77.0
61 43.3 35.2 52.8









#33
 1,2-Dichloroethane-d4
 Concen: 53.883 ug/l
 RT: 5.958 min Scan# 799
 Delta R.T. 0.000 min
 Lab File: VX033259.D
 Acq: 06 Dec 2022 15:08

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1206WBSD02

Tgt Ion: 65 Resp: 73214
 Ion Ratio Lower Upper
 65 100
 67 51.1 0.0 101.8




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 12/07/2022
 Supervised By :Mahesh Dadoda 12/07/2022



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.763 min Scan# 931
 Delta R.T. -0.000 min
 Lab File: VX033259.D
 Acq: 06 Dec 2022 15:08

Tgt Ion:114 Resp: 218668
 Ion Ratio Lower Upper
 114 100
 63 18.6 0.0 41.6
 88 15.6 0.0 33.4





#35

Dibromofluoromethane

Concen: 53.403 ug/l

RT: 5.385 min Scan# 705

Delta R.T. 0.000 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02

Tgt Ion:113 Resp: 6583





Ion	Ratio	Lower	Upper
113	100		
111	102.4	82.4	123.6
192	19.4	15.4	23.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/07/2022

Supervised By :Mahesh Dadoda 12/07/2022



#36

1,1-Dichloropropene

Concen: 18.763 ug/l

RT: 5.696 min Scan# 756




Delta R.T. 0.000 min


Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Tgt Ion: 75 Resp: 36984

Ion	Ratio	Lower	Upper
75	100		
110	36.5	18.4	55.2
77	31.5	24.5	36.7





#37

Ethyl Acetate

Concen: 19.997 ug/l

RT: 4.714 min Scan# 595

Delta R.T. -0.001 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02

Tgt Ion: 43 Resp: 34389





Ion	Ratio	Lower	Upper
43	100		
61	14.6	11.5	17.3
70	11.3	9.1	13.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/07/2022

Supervised By :Mahesh Dadoda 12/07/2022



#38

Carbon Tetrachloride

Concen: 19.401 ug/l

RT: 5.678 min Scan# 753




Delta R.T. -0.006 min


Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Tgt Ion:117 Resp: 42424

Ion	Ratio	Lower	Upper
117	100		
119	93.2	77.2	115.8
121	31.3	25.8	38.6





#39

Methylcyclohexane

Concen: 18.730 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. -0.006 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02

Tgt Ion: 83 Resp: 4212





Ion	Ratio	Lower	Upper
83	100		
55	76.1	58.4	87.6
98	45.8	37.2	55.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/07/2022

Supervised By :Mahesh Dadoda 12/07/2022



#40

Benzene

Concen: 19.515 ug/l

RT: 6.037 min Scan# 812




Delta R.T. 0.000 min


Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Tgt Ion: 78 Resp: 107852

Ion	Ratio	Lower	Upper
78	100		
77	25.0	19.0	28.4






#41
Methacrylonitrile
Concen: 20.976 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02






Tgt Ion: 41 Resp: 2029
Ion Ratio Lower Upper
41 100
39 59.4 47.5 71.3
67 74.1 60.5 90.7
52 31.6 27.0 40.6

Manual Integrations

APPROVED




Reviewed By :Krupa Patel 12/07/2022


Supervised By :Mahesh Dadoda 12/07/2022



#42
1,2-Dichloroethane
Concen: 19.980 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.006 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Tgt Ion: 62 Resp: 44460
Ion Ratio Lower Upper
62 100
98 8.7 0.0 16.0





#43

Isopropyl Acetate

Concen: 20.043 ug/l

RT: 6.342 min Scan# 80

Delta R.T. 0.000 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02

Tgt Ion: 43 Resp: 5700





Ion	Ratio	Lower	Upper
43	100		
61	19.4	15.4	23.0
87	14.5	10.8	16.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/07/2022

Supervised By :Mahesh Dadoda 12/07/2022



#44

Trichloroethene

Concen: 19.468 ug/l

RT: 7.123 min Scan# 990




Delta R.T. -0.006 min


Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Tgt Ion:130 Resp: 30192

Ion	Ratio	Lower	Upper
130	100		
95	98.7	0.0	195.2





#45

1,2-Dichloropropane

Concen: 20.373 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. -0.001 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02

Tgt Ion: 63 Resp: 27820

Ion Ratio Lower Upper

63 100





65 34.4 24.6 36.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/07/2022

Supervised By :Mahesh Dadoda 12/07/2022



#46

Dibromomethane

Concen: 20.267 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.006 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08




Tgt Ion: 93 Resp: 21599


Ion Ratio Lower Upper

93 100

95 81.1 65.8 98.8

174 102.8 79.8 119.6









#47
Bromodichloromethane
Concen: 20.742 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02

Tgt Ion: 83 Resp: 4271
Ion Ratio Lower Upper
83 100
85 67.8 52.2 78.4
127 9.4 7.4 11.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#48
Methyl methacrylate
Concen: 19.811 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Tgt Ion: 41 Resp: 29003
Ion Ratio Lower Upper
41 100
69 84.8 65.1 97.7
39 61.2 51.9 77.9





#49

1,4-Dioxane

Concen: 451.310 ug/l

RT: 7.702 min Scan# 1085

Delta R.T. 0.031 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02





Tgt Ion: 88 Resp: 12162
 Ion Ratio Lower Upper
 88 100
 43 30.8 26.6 39.8
 58 66.3 51.7 77.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/07/2022

Supervised By :Mahesh Dadoda 12/07/2022



#50

Toluene-d8

Concen: 49.507 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. -0.006 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Tgt Ion: 98 Resp: 183710
 Ion Ratio Lower Upper
 98 100
 100 64.7 50.9 76.3









#51
4-Methyl-2-Pentanone
Concen: 104.951 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02

Tgt Ion: 43 Resp: 18418
Ion Ratio Lower Upper
43 100
58 39.7 30.9 46.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#52
Toluene
Concen: 19.649 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Tgt Ion: 92 Resp: 71442
Ion Ratio Lower Upper
92 100
91 174.6 138.3 207.5





#53

t-1,3-Dichloropropene

Concen: 20.193 ug/l

RT: 8.976 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02

Tgt Ion: 75 Resp: 4261

Ion Ratio Lower Upper

75 100





77 29.2 25.1 37.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/07/2022

Supervised By :Mahesh Dadoda 12/07/2022



#54

cis-1,3-Dichloropropene

Concen: 20.441 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08




Tgt Ion: 75 Resp: 45378


Ion Ratio Lower Upper

75 100

77 31.7 25.1 37.7

39 49.9 45.5 68.3









#55
1,1,2-Trichloroethane
Concen: 21.177 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Instrument : MSVOA_X
ClientSampleId : VX1206WBSD02

Tgt Ion	Ratio	Lower	Upper
97	100		
83	79.9	68.1	102.1
85	54.3	43.3	64.9
99	62.4	50.4	75.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#56
Ethyl methacrylate
Concen: 21.020 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Tgt Ion	Ratio	Lower	Upper
69	100		
41	64.4	54.3	81.5
39	38.5	35.2	52.8





#57

1,3-Dichloropropane

Concen: 20.494 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02

Tgt Ion: 76 Resp: 48970

Ion Ratio Lower Upper

76 100





78 32.5 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/07/2022

Supervised By :Mahesh Dadoda 12/07/2022



#58

2-Chloroethyl Vinyl ether

Concen: 105.579 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.006 min

Lab File: VX033259.D




Acq: 06 Dec 2022 15:08


Tgt Ion: 63 Resp: 102257

Ion Ratio Lower Upper

63 100

106 29.7 23.3 34.9









#59
2-Hexanone
Concen: 109.140 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02

Tgt Ion: 43 Resp: 14149
Ion Ratio Lower Upper
43 100
58 54.3 26.5 79.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#60
Dibromochloromethane
Concen: 21.788 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.007 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Tgt Ion:129 Resp: 34725
Ion Ratio Lower Upper
129 100
127 78.2 38.6 115.8





#61
1,2-Dibromoethane
Concen: 20.494 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02





Tgt Ion:107 Resp: 3202
Ion Ratio Lower Upper
107 100
109 97.0 76.0 114.0

Manual Integrations

APPROVED




Reviewed By :Krupa Patel 12/07/2022


Supervised By :Mahesh Dadoda 12/07/2022



#62
4-Bromofluorobenzene
Concen: 53.047 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08


Tgt Ion: 95 Resp: 85903
Ion Ratio Lower Upper
95 100
174 80.9 0.0 164.0
176 79.2 0.0 154.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02



Tgt Ion:117 Resp: 197470
Ion Ratio Lower Upper
117 100
82 54.9 44.3 66.5
119 31.9 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#64
Tetrachloroethene
Concen: 19.263 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Tgt Ion:164 Resp: 26926
Ion Ratio Lower Upper
164 100
166 130.2 105.4 158.2
129 96.5 78.2 117.2
131 92.4 72.5 108.7









#65
Chlorobenzene
Concen: 20.531 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02

Tgt Ion:112 Resp: 8083
Ion Ratio Lower Upper
112 100
114 32.3 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 21.890 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08


Tgt Ion:131 Resp: 31579
Ion Ratio Lower Upper
131 100
133 92.7 47.6 142.9
119 62.9 32.4 97.0





#67
Ethyl Benzene
Concen: 20.154 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02




Tgt Ion: 91 Resp: 140754
Ion Ratio Lower Upper
91 100
106 31.8 25.0 37.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022




#68
m/p-Xylenes
Concen: 40.433 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.006 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08



Tgt Ion:106 Resp: 110578
Ion Ratio Lower Upper
106 100
91 203.7 164.5 246.7









#69
o-Xylene
Concen: 20.498 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02

Tgt Ion:106 Resp: 5424
Ion Ratio Lower Upper
106 100
91 219.1 109.2 327.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#70
Styrene
Concen: 20.931 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.007 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Tgt Ion:104 Resp: 90613
Ion Ratio Lower Upper
104 100
78 50.8 41.8 62.8
103 54.8 44.5 66.7





#71

Bromoform

Concen: 22.698 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02





Tgt Ion:173 Resp: 2510
 Ion Ratio Lower Upper
 173 100
 175 48.6 23.9 71.8
 254 0.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/07/2022

Supervised By :Mahesh Dadoda 12/07/2022



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.024 min Scan# 1794


Delta R.T. 0.000 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08


Tgt Ion:152 Resp: 101506
 Ion Ratio Lower Upper
 152 100
 115 70.2 44.0 132.0
 150 162.3 0.0 351.8





#73
Isopropylbenzene
Concen: 20.030 ug/l
RT: 10.963 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02




Tgt Ion:105 Resp: 145285
Ion Ratio Lower Upper
105 100
120 26.5 13.1 39.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022




#74
N-ethyl acetate
Concen: 19.882 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Tgt Ion: 43 Resp: 49059
Ion Ratio Lower Upper
43 100
70 44.6 33.6 50.4
55 26.3 19.4 29.2
61 25.2 18.8 28.2





#75
1,1,2,2-Tetrachloroethane
Concen: 21.055 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02




Tgt Ion: 83 Resp: 46095
Ion Ratio Lower Upper
83 100
131 11.1 5.4 16.2
85 65.6 31.9 95.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022




#76
1,2,3-Trichloropropane
Concen: 21.197 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.006 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Tgt Ion: 75 Resp: 40749
Ion Ratio Lower Upper
75 100
77 39.3 20.0 60.0





#77
Bromobenzene
Concen: 20.224 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.006 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02




Tgt Ion:156 Resp: 36078
Ion Ratio Lower Upper
156 100
77 150.9 76.1 228.4
158 97.4 48.8 146.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022




#78
n-propylbenzene
Concen: 19.946 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08



Tgt Ion: 91 Resp: 165419
Ion Ratio Lower Upper
91 100
120 23.5 11.6 34.8









#79
2-Chlorotoluene
Concen: 19.768 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02

Tgt Ion: 91 Resp: 99572
Ion Ratio Lower Upper
91 100
126 34.9 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#80
1,3,5-Trimethylbenzene
Concen: 20.169 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08


Tgt Ion:105 Resp: 122896
Ion Ratio Lower Upper
105 100
120 47.4 23.6 71.0





#81
trans-1,4-Dichloro-2-butene
Concen: 20.722 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02




Tgt Ion: 75 Resp: 1225
Ion Ratio Lower Upper
75 100
53 105.2 86.9 130.3
89 47.7 38.3 57.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022




#82
4-Chlorotoluene
Concen: 19.830 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Tgt Ion: 91 Resp: 117051
Ion Ratio Lower Upper
91 100
126 30.3 14.8 44.4





#83
tert-Butylbenzene
Concen: 20.496 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02




Tgt Ion:119 Resp: 12432
Ion Ratio Lower Upper
119 100
91 57.2 29.4 88.2
134 22.7 11.0 33.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022




#84
1,2,4-Trimethylbenzene
Concen: 20.480 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08



Tgt Ion:105 Resp: 124432
Ion Ratio Lower Upper
105 100
120 43.7 21.9 65.5









#85
 sec-Butylbenzene
 Concen: 20.281 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX033259.D
 Acq: 06 Dec 2022 15:08

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1206WBSD02

Tgt Ion:105 Resp: 147168
 Ion Ratio Lower Upper
 105 100
 134 20.0 9.9 29.7




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 12/07/2022
 Supervised By :Mahesh Dadoda 12/07/2022



#86
 p-Isopropyltoluene
 Concen: 20.686 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX033259.D
 Acq: 06 Dec 2022 15:08

Tgt Ion:119 Resp: 127180
 Ion Ratio Lower Upper
 119 100
 134 26.0 12.9 38.7
 91 24.6 12.5 37.5





#87

1,3-Dichlorobenzene

Concen: 20.455 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08

Instrument :

MSVOA_X

ClientSampleId :

VX1206WBSD02





Tgt Ion:146 Resp: 68440
 Ion Ratio Lower Upper
 146 100
 111 40.1 20.6 61.8
 148 62.8 31.7 95.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 12/07/2022

Supervised By :Mahesh Dadoda 12/07/2022



#88

1,4-Dichlorobenzene

Concen: 20.149 ug/l




RT: 12.042 min Scan# 1797


Delta R.T. 0.000 min

Lab File: VX033259.D

Acq: 06 Dec 2022 15:08


Tgt Ion:146 Resp: 69045
 Ion Ratio Lower Upper
 146 100
 111 40.8 20.5 61.6
 148 63.9 32.2 96.6





#89
n-Butylbenzene
Concen: 20.046 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02



Tgt Ion: 91 Resp: 106174
Ion Ratio Lower Upper
91 100
92 52.0 26.2 78.6
134 26.1 12.9 38.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#90
Hexachloroethane
Concen: 21.026 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.006 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08


Tgt Ion:117 Resp: 21238
Ion Ratio Lower Upper
117 100
201 75.8 37.6 112.9





#91
1,2-Dichlorobenzene
Concen: 20.785 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02




Tgt Ion:146 Resp: 6765
Ion Ratio Lower Upper
146 100
111 41.5 21.5 64.5
148 64.4 32.0 96.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 22.667 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08



Tgt Ion: 75 Resp: 9901
Ion Ratio Lower Upper
75 100
155 85.1 41.6 124.8
157 110.0 54.1 162.4









#93
1,2,4-Trichlorobenzene
Concen: 20.930 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08

Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02

Tgt Ion:180 Resp: 4199
Ion Ratio Lower Upper
180 100
182 94.7 48.4 145.0
145 30.9 16.2 48.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#94
Hexachlorobutadiene
Concen: 19.680 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08


Tgt Ion:225 Resp: 17420
Ion Ratio Lower Upper
225 100
223 63.5 31.9 95.9
227 62.8 31.9 95.5





#95
Naphthalene
Concen: 21.367 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.006 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08




Instrument :
MSVOA_X
ClientSampleId :
VX1206WBSD02




Tgt Ion:128 Resp: 13868
Ion Ratio Lower Upper
128 100
127 12.7 10.2 15.4
129 10.7 8.9 13.3

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 12/07/2022
Supervised By :Mahesh Dadoda 12/07/2022



#96
1,2,3-Trichlorobenzene
Concen: 21.207 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX033259.D
Acq: 06 Dec 2022 15:08



Tgt Ion:180 Resp: 42668
Ion Ratio Lower Upper
180 100
182 93.9 47.5 142.6
145 33.4 16.9 50.6

