



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120622\
 Data File : VX033253.D
 Acq On : 06 Dec 2022 08:22
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Title : SW846 8260
 Last Update : Tue Nov 22 13:58:46 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	7817	PASS
75	95	30	60	52.0	21549	PASS
95	95	100	100	100.0	41461	PASS
96	95	5	9	6.2	2560	PASS
173	174	0.00	2	0.8	262	PASS
174	95	50	100	79.0	32768	PASS
175	174	5	9	7.4	2441	PASS
176	174	95	101	97.9	32069	PASS
177	176	5	9	6.9	2206	PASS