



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120623\
Data File : VX039084.D
Acq On : 06 Dec 2023 08:38
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Mon Nov 27 03:22:25 2023



AutoFind: Averaged scan 1634 to 1649; Bkg corrected with scan 1633

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.0	2334	PASS
75	95	30	60	52.8	6156	PASS
95	95	100	100	100.0	11669	PASS
96	95	5	9	6.4	742	PASS
173	174	0.00	2	0.5	38	PASS
174	95	50	100	66.1	7714	PASS
175	174	5	9	7.4	568	PASS
176	174	95	101	96.9	7474	PASS
177	176	5	9	6.2	467	PASS