

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120718\
 Data File : VX006342.D
 Acq On : 07 Dec 2018 12:58
 Operator : JC/MD
 Sample : J6259-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW305I-20181205

Quant Time: Dec 08 00:57:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	860124	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1309873	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1183449	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	584514	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	429052	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
35) Dibromofluoromethane	5.51	113	404679	47.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.44%	
50) Toluene-d8	8.71	98	1512441	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
62) 4-Bromofluorobenzene	11.13	95	564267	48.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.08%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	7742	0.952	ug/l	99
24) 1,1-Dichloroethane	3.70	63	20621	1.502	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	29031	3.138	ug/l	89
44) Trichloroethene	7.22	130	8297865	829.090	ug/l	95
49) 1,4-Dioxane	7.76	88	3425	13.870	ug/l #	80
64) Tetrachloroethene	9.34	164	18113	2.032	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120718\
Data File : VX006342.D
Acq On : 07 Dec 2018 12:58
Operator : JC/MD
Sample : J6259-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW305I-20181205

Quant Time: Dec 08 00:57:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

