

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120718\
 Data File : VX006347.D
 Acq On : 07 Dec 2018 15:11
 Operator : JC/MD
 Sample : J6252-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CPA-CPMW11D-M16W4-11-28-18

Quant Time: Dec 10 10:00:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	814072	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1247781	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1130527	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	547198	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	406692	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
35) Dibromofluoromethane	5.51	113	388263	48.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.12%	
50) Toluene-d8	8.71	98	1449827	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
62) 4-Bromofluorobenzene	11.13	95	540907	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
Target Compounds						
44) Trichloroethene	7.21	130	3639	0.382	ug/l	100
64) Tetrachloroethene	9.34	164	29787	3.499	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120718\
Data File : VX006347.D
Acq On : 07 Dec 2018 15:11
Operator : JC/MD
Sample : J6252-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CPA-CPMW11D-M16W4-11-28-18

Quant Time: Dec 10 10:00:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

