

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX120718\
 Data File : VX006351.D
 Acq On : 07 Dec 2018 16:58
 Operator : JC/MD
 Sample : J6252-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CPA-CPMW10S-M16W4-11-27-18

Quant Time: Dec 10 10:04:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	833842	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1239879	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1117964	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	556249	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	403751	47.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.86%	
35) Dibromofluoromethane	5.50	113	387498	48.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.54%	
50) Toluene-d8	8.71	98	1429908	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
62) 4-Bromofluorobenzene	11.13	95	533037	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
Target Compounds						
30) Chloroform	5.21	83	14062	1.006	ug/l	92
44) Trichloroethene	7.21	130	4896	0.517	ug/l	85
64) Tetrachloroethene	9.34	164	109807	13.042	ug/l	96
84) 1,2,4-Trimethylbenzene	11.80	105	47617	1.443	ug/l	100
93) 1,2,4-Trichlorobenzene	13.65	180	6452	0.470	ug/l	95
95) Naphthalene	13.83	128	19696	0.461	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX120718\
Data File : VX006351.D
Acq On : 07 Dec 2018 16:58
Operator : JC/MD
Sample : J6252-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CPA-CPMW10S-M16W4-11-27-18

Quant Time: Dec 10 10:04:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

