

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120721\
 Data File : VX025567.D
 Acq On : 07 Dec 2021 10:11
 Operator : JC/MD
 Sample : VX1207MBL01
 Misc : 5.0g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK643

Quant Time: Dec 08 05:32:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 08 05:26:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	143043	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	130174	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	60549	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	54066	55.384	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	110.760%
7) Chloroethane-d5	1.666	69	43409	53.878	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	107.760%
11) 1,1-Dichloroethene-d2	2.306	63	72436	46.537	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.080%
21) 2-Butanone-d5	4.452	46	72917	100.342	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.340%
24) Chloroform-d	5.056	84	90951	53.309	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.620%
26) 1,2-Dichloroethane-d4	5.958	65	58943	55.426	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.860%
32) Benzene-d6	5.977	84	193560	56.104	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.200%
36) 1,2-Dichloropropane-d6	7.306	67	56248	52.658	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.320%
41) Toluene-d8	8.647	98	178754	54.893	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.780%
43) trans-1,3-Dichloroprop...	8.952	79	25152	46.835	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.680%
47) 2-Hexanone-d5	9.384	63	53309	95.875	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.880%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	76494	50.671	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.340%
66) 1,2-Dichlorobenzene-d4	12.323	152	65485	55.142	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.280%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120721\
Data File : VX025567.D
Acq On : 07 Dec 2021 10:11
Operator : JC/MD
Sample : VX1207MBL01
Misc : 5.0g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK643

Quant Time: Dec 08 05:32:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 08 05:26:42 2021
Response via : Initial Calibration

