

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120721\
 Data File : VX025594.D
 Acq On : 07 Dec 2021 22:21
 Operator : JC/MD
 Sample : M4886-09ME
 Misc : 4.40g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EX890ME

Quant Time: Dec 08 13:55:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 08 07:24:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	142387	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.061	117	126943	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	63590	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	46827	48.189	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.380%
7) Chloroethane-d5	1.648	69	23130	28.840	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	57.680%#
11) 1,1-Dichloroethene-d2	2.282	63	65101	42.018	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.040%
21) 2-Butanone-d5	4.489	46	62500	86.403	ug/L	0.03
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.400%
24) Chloroform-d	5.068	84	80604	47.462	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.920%
26) 1,2-Dichloroethane-d4	5.958	65	52927	49.999	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.000%
32) Benzene-d6	5.971	84	179670	53.403	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.800%
36) 1,2-Dichloropropane-d6	7.312	67	52152	50.067	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.140%
41) Toluene-d8	8.653	98	164574	51.824	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.640%
43) trans-1,3-Dichloroprop...	8.958	79	20774	39.668	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.340%
47) 2-Hexanone-d5	9.397	63	45751	84.377	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.380%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	67390	45.777	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.560%
66) 1,2-Dichlorobenzene-d4	12.323	152	63811	51.163	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120721\
Data File : VX025594.D
Acq On : 07 Dec 2021 22:21
Operator : JC/MD
Sample : M4886-09ME
Misc : 4.40g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EX890ME

Quant Time: Dec 08 13:55:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 08 07:24:50 2021
Response via : Initial Calibration

