

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120821\
 Data File : VX025624.D
 Acq On : 08 Dec 2021 16:23
 Operator : JC/MD
 Sample : M4883-08ME
 Misc : 5.06g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EW9G4ME

Quant Time: Dec 09 01:43:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXMLM112221WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 09 01:41:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	143679	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.061	117	130352	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	66015	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	43543	44.407	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery =	88.820%		
7) Chloroethane-d5	1.648	69	23041	28.471	ug/L	-0.02
Spiked Amount	50.000	Range 70 - 130	Recovery =	56.940%#		
11) 1,1-Dichloroethene-d2	2.282	63	63358	40.525	ug/L	-0.02
Spiked Amount	50.000	Range 60 - 125	Recovery =	81.040%		
21) 2-Butanone-d5	4.489	46	66447	91.034	ug/L	0.04
Spiked Amount	100.000	Range 40 - 130	Recovery =	91.030%		
24) Chloroform-d	5.068	84	81335	47.462	ug/L	0.01
Spiked Amount	50.000	Range 70 - 125	Recovery =	94.920%		
26) 1,2-Dichloroethane-d4	5.958	65	54284	50.819	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	101.640%		
32) Benzene-d6	5.977	84	181677	52.588	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	105.180%		
36) 1,2-Dichloropropane-d6	7.312	67	53510	50.027	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery =	100.060%		
41) Toluene-d8	8.653	98	166559	51.078	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery =	102.160%		
43) trans-1,3-Dichloroprop...	8.958	79	20870	38.809	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery =	77.620%		
47) 2-Hexanone-d5	9.397	63	47826	85.897	ug/L	0.01
Spiked Amount	100.000	Range 45 - 130	Recovery =	85.900%		
56) 1,1,2,2-Tetrachloroeth...	11.195	84	70301	46.505	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery =	93.020%		
66) 1,2-Dichlorobenzene-d4	12.323	152	66891	51.662	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery =	103.320%		
Target Compounds						
34) Trichloroethene	7.123	95	47297	51.638	ug/L	96
35) Methylcyclohexane	7.373	83	1196	0.830	ug/L #	84
42) Toluene	8.720	91	5313	1.383	ug/L	94
46) Tetrachloroethene	9.275	164	1599	2.059	ug/L	89
53) m,p-Xylene	10.305	106	1499	0.897	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120821\
 Data File : VX025624.D
 Acq On : 08 Dec 2021 16:23
 Operator : JC/MD
 Sample : M4883-08ME
 Misc : 5.06g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EW9G4ME

Quant Time: Dec 09 01:43:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 09 01:41:02 2021
 Response via : Initial Calibration

