

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120821\
 Data File : VX025638.D
 Acq On : 08 Dec 2021 22:36
 Operator : JC/MD
 Sample : M4975-03
 Misc : 6.54g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BGKS7

Quant Time: Dec 09 03:32:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 09 03:12:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	262346	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.061	117	233040	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	117565	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.373	65	51686	28.869	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	57.740%#
7) Chloroethane-d5	1.648	69	27387	18.534	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	37.060%#
11) 1,1-Dichloroethene-d2	2.282	63	74586	26.127	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	52.260%#
21) 2-Butanone-d5	4.489	46	79114	59.361	ug/L	0.04
Spiked Amount	100.000	Range	40 - 130	Recovery	=	59.360%
24) Chloroform-d	5.068	84	99706	31.864	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	63.720%#
26) 1,2-Dichloroethane-d4	5.964	65	67081	34.393	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	68.780%#
32) Benzene-d6	5.970	84	216408	35.039	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	70.080%
36) 1,2-Dichloropropane-d6	7.311	67	64289	33.620	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	67.240%#
41) Toluene-d8	8.653	98	198244	34.006	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	68.020%#
43) trans-1,3-Dichloroprop...	8.957	79	25899	26.939	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	53.880%#
47) 2-Hexanone-d5	9.396	63	57186	57.450	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	57.450%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	84889	31.411	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	62.820%#
66) 1,2-Dichlorobenzene-d4	12.323	152	77838	33.757	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	67.520%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120821\
Data File : VX025638.D
Acq On : 08 Dec 2021 22:36
Operator : JC/MD
Sample : M4975-03
Misc : 6.54g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BGKS7

Quant Time: Dec 09 03:32:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 09 03:12:30 2021
Response via : Initial Calibration

